



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB101685

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	03/28/2019	08:05
2	CAL2	1	Water	NA	NA	20.1	7.01	03/28/2019	08:06
3	CAL3	1	Water	NA	NA	20.1	10.02	03/28/2019	08:07
4	ICV	1	Water	NA	NA	20.1	6.99	03/28/2019	08:09
5	CCV1	1	Water	NA	NA	20.1	2.01	03/28/2019	08:10
6	K2124-01	1	Water	NA	NA	17.0	6.56	03/28/2019	08:11
7	K2124-01DUP	1	Water	NA	NA	17.1	6.57	03/28/2019	08:12
8	K2124-04	1	Water	NA	NA	17.0	6.78	03/28/2019	08:13
9	CCV2	1	Water	NA	NA	20.2	12.01	03/28/2019	08:15

WORKLIST(Hardcopy Internal Chain)

2B 101683

WorkList Name : ph k2124

WorkList ID : 124046

Date : 03-28-2019 07:55:35

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/27/2019	Water	K2124-01 -A	pH	Cool 4 deg C	PSEG01	K31	SAMPLE-1	03/27/2019	9040C
03/27/2019	Water	K2124-04 -A	pH	Cool 4 deg C	PSEG01	K31	SAMPLE-2	03/27/2019	9040C

Date/Time 03-28-19 8:00
 Received by: [Signature]
 Relinquished by: [Signature]

Date/Time 03-28-19 9:15
 Received by: [Signature]
 Relinquished by: [Signature]