


LB101691

7470A

INSTRUMENT ID: C01

 μ Abs.:

42794



Linear

A= 0.0000e+000

B= 2.3105e-004

C= 5.2046e-003

Rho= 0.9993976

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.037	0.037	139	0.000	139				
0.2	0.200	0.215	0.015	906	0.0 %	906				
2.5	2.500	2.467	-0.033	10655	0.0 %	10655				
5.0	5.000	4.844	-0.156	20941	0.0 %	20941				
7.5	7.500	7.745	0.245	33496	0.0 %	33496				
10.0	10.000	9.893	-0.107	42794	0.0 %	42794				

LB101691 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 28 Mar 2019 11:01:56

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	139	-	0.00007470A	PPB		28 Mar 2019 11:04:16	S	Std
0.2 - 1	50.2	906	-	0.20007470A	PPB		28 Mar 2019 11:06:25	S	Std
2.5 - 1	52.5	10655	-	2.50007470A	PPB		28 Mar 2019 11:08:33	S	Std
5.0 - 1	55.0	20941	-	5.00007470A	PPB		28 Mar 2019 11:10:41	S	Std
7.5 - 1	57.5	33496	-	7.50007470A	PPB		28 Mar 2019 11:12:54	S	Std
10.0 - 1	510.0	42794	-	10.00007470A	PPB		28 Mar 2019 11:15:01	S	Std
ICV73 - 1	ICV73	17149	3.9675	-7470A	PPB		28 Mar 2019 11:19:09	U	SMPL
ICB73 - 1	ICB73	77	0.0230	-7470A	PPB		28 Mar 2019 11:21:15	U	SMPL
CCV68 - 1	CCV68	21263	4.9181	-7470A	PPB		28 Mar 2019 11:23:20	U	SMPL
CCB68 - 1	CCB68	-9	0.0031	-7470A	PPB		28 Mar 2019 11:25:26	U	SMPL
CRA - 1	CRA	872	0.2067	-7470A	PPB		28 Mar 2019 11:27:33	U	SMPL
HighStd - 1	HighStd	43393	10.0313	-7470A	PPB		28 Mar 2019 11:29:39	U	SMPL
ChkStd - 1	ChkStd	30837	7.1302	-7470A	PPB		28 Mar 2019 11:31:45	U	SMPL
PB118354BL - 1	PBW	48	0.0163	-7470A	PPB		28 Mar 2019 11:33:53	U	SMPL
PB118354BS - 1	LCSW	16779	3.8820	-7470A	PPB		28 Mar 2019 11:36:01	U	SMPL
K2086-08 - 1	CS-11	165	0.0433	-7470A	PPB		28 Mar 2019 11:44:12	U	SMPL
K2086-08DUP - 1	CS-11DUP	161	0.0424	-7470A	PPB		28 Mar 2019 11:46:19	U	SMPL
K2086-08MS - 1	CS-11MS	16884	3.9063	-7470A	PPB		28 Mar 2019 11:48:27	U	SMPL
K2086-08MSD - 1	CS-11MSD	17119	3.9606	-7470A	PPB		28 Mar 2019 11:50:35	U	SMPL
K2086-09 - 1	CS-12	80	0.0237	-7470A	PPB		28 Mar 2019 11:52:41	U	SMPL
CCV69 - 1	CCV69	20915	4.8377	-7470A	PPB		28 Mar 2019 11:54:47	U	SMPL
CCB69 - 1	CCB69	-11	0.0027	-7470A	PPB		28 Mar 2019 11:56:53	U	SMPL
K2086-10 - 1	CS-13	148	0.0394	-7470A	PPB		28 Mar 2019 11:59:00	U	SMPL
K2086-11 - 1	CS-14	152	0.0403	-7470A	PPB		28 Mar 2019 12:01:06	U	SMPL
K2086-12 - 1	CS-15	149	0.0396	-7470A	PPB		28 Mar 2019 12:03:13	U	SMPL
K2086-13 - 1	CS-16	144	0.0385	-7470A	PPB		28 Mar 2019 12:05:19	U	SMPL
K2086-14 - 1	CS-18	135	0.0364	-7470A	PPB		28 Mar 2019 12:07:26	U	SMPL
K2090-03 - 1	P-5(0-10)	179	0.0466	-7470A	PPB		28 Mar 2019 12:09:33	U	SMPL
K2097-07 - 1	CS-20	136	0.0366	-7470A	PPB		28 Mar 2019 12:11:41	U	SMPL
K2097-08 - 1	CS-21	145	0.0387	-7470A	PPB		28 Mar 2019 12:13:49	U	SMPL
K2097-09 - 1	CS-22	160	0.0422	-7470A	PPB		28 Mar 2019 12:15:56	U	SMPL
K2097-10 - 1	CS-19	163	0.0429	-7470A	PPB		28 Mar 2019 12:18:03	U	SMPL
CCV70 - 1	CCV70	20794	4.8097	-7470A	PPB		28 Mar 2019 12:20:09	U	SMPL
CCB70 - 1	CCB70	5	0.0064	-7470A	PPB		28 Mar 2019 12:22:16	U	SMPL
K2097-11 - 1	CS-17	148	0.0394	-7470A	PPB		28 Mar 2019 12:24:22	U	SMPL
K2097-12 - 1	CS-6	162	0.0426	-7470A	PPB		28 Mar 2019 12:26:29	U	SMPL
K2105-04 - 1	NJ-WATER	161	0.0424	-7470A	PPB		28 Mar 2019 12:28:35	U	SMPL
PB118355BL - 1	PBW	156	0.0412	-7470A	PPB		28 Mar 2019 12:30:42	U	SMPL
PB118355BS - 1	LCSW	17391	4.0234	-7470A	PPB		28 Mar 2019 12:32:48	U	SMPL
K2093-17 - 1	30092-30093-5	437	0.1062	-7470A	PPB		28 Mar 2019 12:34:55	U	SMPL
K2093-17DUP - 1	30092-30093-5DUP	344	0.0847	-7470A	PPB		28 Mar 2019 12:37:02	U	SMPL
K2093-17MS - 1	30092-30093-5MS	19575	4.5281	-7470A	PPB		28 Mar 2019 12:44:51	U	SMPL
K2093-17MSD - 1	30092-30093-5MSD	19296	4.4636	-7470A	PPB		28 Mar 2019 12:46:58	U	SMPL
K2093-20 - 1	30099-30100-OW-6	803	0.1907	-7470A	PPB		28 Mar 2019 12:49:04	U	SMPL
CCV71 - 1	CCV71	21983	5.0844	-7470A	PPB		28 Mar 2019 12:51:11	U	SMPL
CCB71 - 1	CCB71	-27	-0.0010	-7470A	PPB		28 Mar 2019 12:53:17	U	SMPL
K2121-01 - 1	MW-11	406	0.0990	-7470A	PPB		28 Mar 2019 12:55:24	U	SMPL
K2121-02 - 1	TW-1	1365	0.3206	-7470A	PPB		28 Mar 2019 12:57:31	U	SMPL
K2121-03 - 1	MW-9	281	0.0701	-7470A	PPB		28 Mar 2019 12:59:37	U	SMPL
K2121-04 - 1	MW-8	791	0.1880	-7470A	PPB		28 Mar 2019 13:01:44	U	SMPL
K2121-05 - 1	MW-7	418	0.1018	-7470A	PPB		28 Mar 2019 13:03:51	U	SMPL
K2121-06 - 1	MW-6	332	0.0819	-7470A	PPB		28 Mar 2019 13:05:58	U	SMPL
K2121-07 - 1	FIELD-BLANK	224	0.0570	-7470A	PPB		28 Mar 2019 13:08:05	U	SMPL
K2121-08 - 1	DUP	409	0.0997	-7470A	PPB		28 Mar 2019 13:10:13	U	SMPL
K2124-01 - 1	SAMPLE-1	53	0.0175	-7470A	PPB		28 Mar 2019 13:12:22	U	SMPL
K2124-04 - 1	SAMPLE-2	5	0.0064	-7470A	PPB		28 Mar 2019 13:14:28	U	SMPL
CCV72 - 1	CCV72	21705	5.0202	-7470A	PPB		28 Mar 2019 13:16:35	U	SMPL
CCB72 - 1	CCB72	18	0.0094	-7470A	PPB		28 Mar 2019 13:18:41	U	SMPL
PB118354TB - 1	PB118298TB	212	0.0542	-7470A	PPB		28 Mar 2019 13:20:48	U	SMPL
K2086-08LX5 - 1		11	0.0077	-7470A	PPB		28 Mar 2019 13:22:54	U	SMPL
K2086-08A - 1		17488	4.0458	-7470A	PPB		28 Mar 2019 13:25:01	U	SMPL
K2093-17LX5 - 1		-1	0.0050	-7470A	PPB		28 Mar 2019 13:27:07	U	SMPL
K2093-17A - 1		20330	4.7025	-7470A	PPB		28 Mar 2019 13:29:14	U	SMPL
CCV73 - 1	CCV73	21739	5.0281	-7470A	PPB		28 Mar 2019 13:31:21	U	SMPL
CCB73 - 1	CCB73	-27	-0.0010	-7470A	PPB		28 Mar 2019 13:33:28	U	SMPL