



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB101694

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	03/28/2019	12:35
2	CAL2	1	Water	NA	NA	20.1	7.00	03/28/2019	12:36
3	CAL3	1	Water	NA	NA	20.1	10.02	03/28/2019	12:37
4	ICV	1	Water	NA	NA	20.1	7.01	03/28/2019	12:39
5	CCV1	1	Water	NA	NA	20.1	2.01	03/28/2019	12:40
6	K2145-01	1	Water	NA	NA	19.0	5.44	03/28/2019	12:41
7	K2145-01DUP	1	Water	NA	NA	19.1	5.45	03/28/2019	12:42
8	CCV2	1	Water	NA	NA	20.2	12.02	03/28/2019	12:45

18101694

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K2145 WorkList ID : 124083 Date : 03-28-2019 12:24:54

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/27/2019	Water	K2145-01-B	pH	Cool 4 deg C	STER01	F52	POOL-WATER	03/27/2019	SM 4500-H B

Date/Time 03-28-19 12:30
Received by: JS
Relinquished by: CB

Date/Time 03-28-19 1:00
Received by: CB
Relinquished by: JS