



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB101700

Ambient Barometric Pressure (mmHg): 760

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.0	80.6	80.6	03/28/2019	09:45
2	K2124-02		1	7	>100.0	>212.0	>212.0	03/28/2019	10:20
3	K2124-02DUP		1	7	>100.0	>212.0	>212.0	03/28/2019	10:50
4	K2124-05		1	8	>100.0	>212.0	>212.0	03/28/2019	11:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 101700

WorkList Name : FP32819

WorkList ID : 124063

Date : 03-28-2019 09:15:46

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/01/2019	Water	K2124-02	Flash Point	Cool 4 deg C	PSEG01	K31	SAMPLE-1	03/27/2019	1010A
04/01/2019	Water	K2124-05	Flash Point	Cool 4 deg C	PSEG01	K31	SAMPLE-2	03/27/2019	1010A

Date/Time 03/28/19 09:20
 Received by: KS
 Relinquished by: CP

Date/Time 03/28/19 11:25
 Received by: CP
 Relinquished by: KS