



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 3/29/2019

OVENTEMP IN Celsius(°C): 108
Time IN: 17:25
In Date: 03/28/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:11
Out Date: 03/29/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-1
Thermometer ID: %SOLIDS-OVEN

QC:LB101707

Lab ID	Client Sample ID	Dish#	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish + Dry Sample Wt(g) (C)	% Solid
K1685-09	AU-06-032819-A	1	1.13	9.59	8.46	86.6
K1685-11	AU-06-032819-B	2	1.12	9.96	9.22	91.6
K1685-13	AU-06-032819-C	3	1.13	9.98	8.83	87.0
K1698-03	TR-05-032819	4	1.13	9.96	9.27	92.2
K2138-01	S6A (2-10) COMP	5	1.12	9.77	9.28	94.3
K2138-03	S6A (12-20) COMP	6	1.14	9.77	9.22	93.6
K2138-05	S6A/S6B (21-25) COMP	7	1.13	9.79	9.14	92.5
K2138-07	S6B (2-10) COMP	8	1.11	9.86	9.44	95.2
K2138-09	S6B (12-10) COMP	9	1.1	9.72	9.3	95.1
K2138-11	S6A (10)	10	1.14	9.6	9.2	95.3
K2138-13	S6A (20)	11	1.13	9.9	9.45	94.9
K2138-15	S6B (25)	12	1.14	9.83	8.77	87.8
K2138-17	S6B (8)	13	1.15	9.69	9.21	94.4
K2138-19	S6B (22)	14	1.14	9.62	9.12	94.1
K2139-01	S7B (2-10) COMP	15	1.13	9.93	9.42	94.2
K2139-03	S7B (12-20) COMP	16	1.13	9.74	9.21	93.8
K2139-05	S7B (8)	17	1.15	9.72	9.08	92.5
K2139-07	S7B (14)	18	1.13	9.5	8.88	92.6
K2140-01	390321143	19	1.12	9.9	2.99	21.3
K2141-01	390322022	20	1.12	9.85	1.37	2.9
K2142-01	390320161	21	1.13	9.98	1.46	3.7
K2143-01	390320712	22	1.13	9.97	3.2	23.4
K2144-01	390321142	23	1.13	9.69	1.47	4.0
K2153-01	CONC-1	24	1.00	2.00	2.00	100.0
K2153-03	CONC-2	25	1.00	2.00	2.00	100.0
K2153-05	CONC-3	26	1.00	2.00	2.00	100.0
K2153-07	CONC-4	27	1.00	2.00	2.00	100.0
K2153-09	CONC-5	28	1.00	2.00	2.00	100.0
K2153-11	CONC-6	29	1.00	2.00	2.00	100.0
K2153-13	CONC-7	30	1.00	2.00	2.00	100.0
K2153-15	CONC-8	31	1.00	2.00	2.00	100.0
K2153-17	CONC-9	32	1.00	2.00	2.00	100.0
K2153-19	CONC-10	33	1.00	2.00	2.00	100.0
K2153-21	SP-1	34	1.15	9.95	9.14	90.8
K2153-22	SP-1-A	35	1.15	9.95	9.14	90.8
K2153-23	SP-1-B	36	1.15	9.95	9.14	90.8
K2153-24	SP-1	37	1.15	9.95	9.14	90.8
K2153-25	SP-2	38	1.14	9.91	9.12	91.0
K2153-26	SP-2-A	39	1.14	9.91	9.12	91.0



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 3/29/2019

OVENTEMP IN Celsius(°C): 108
Time IN: 17:25
In Date: 03/28/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:11
Out Date: 03/29/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-1
Thermometer ID: %SOLIDS-OVEN

QC:LB101707

Lab ID	Client Sample ID	Dish#	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish + Dry Sample Wt(g) (C)	% Solid
K2153-27	SP-2-B	40	1.14	9.91	9.12	91.0
K2153-28	SP-2	41	1.14	9.91	9.12	91.0
K2155-01	SSSI-0328-S-DH-5-T	42	1.17	9.96	9.13	90.6
K2155-02	SSSI-0328-S-DH-5-T	43	1.19	9.83	8.98	90.2
K2155-03	SSSI-0328-S-DH-5-B	44	1.18	9.88	8.78	87.4
K2155-04	SSSI-0328-S-DH-5-B	45	1.13	9.81	8.7	87.2

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$