


LB101710

INSTRUMENT ID: C01

245.1

 μ Abs.:

41651



Linear

A= 0.0000e+000

B= 2.3597e-004

C= 2.0871e-002

Rho= 0.9995797

Accept = Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.021	0.021	0	0.000	0				
0.05	0.050	0.058	0.008	156	0.0 %	156				
0.2	0.200	0.190	-0.010	715	0.0 %	715				
2.5	2.500	2.420	-0.080	10168	0.0 %	10168				
5.0	5.000	4.986	-0.014	21042	0.0 %	21042				
7.5	7.500	7.726	0.226	32654	0.0 %	32654				
10.0	10.000	9.849	-0.151	41651	0.0 %	41651				

LB101710 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 28 Mar 2019 20:21:25

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	0	-	0.0000	245.1	PPB	28 Mar 2019 20:32:05	S	Std
0.05 - 1	50.05	156	-	0.0500	245.1	PPB	28 Mar 2019 20:34:22	S	Std
0.2 - 1	50.2	715	-	0.2000	245.1	PPB	28 Mar 2019 20:36:39	S	Std
2.5 - 1	52.5	10168	-	2.5000	245.1	PPB	28 Mar 2019 20:38:57	S	Std
5.0 - 1	55.0	21042	-	5.0000	245.1	PPB	28 Mar 2019 20:41:13	S	Std
7.5 - 1	57.5	32654	-	7.5000	245.1	PPB	28 Mar 2019 20:43:29	S	Std
10.0 - 1	510.0	41651	-	10.0000	245.1	PPB	28 Mar 2019 20:45:45	S	Std
ICV76 - 1	ICV76	17631	4.1813	-245.1		PPB	28 Mar 2019 20:53:02	U	SMPL
ICB76 - 1	ICB76	24	0.0265	-245.1		PPB	28 Mar 2019 20:55:16	U	SMPL
CCV83 - 1	CCV83	21218	5.0277	-245.1		PPB	28 Mar 2019 20:57:32	U	SMPL
CCB83 - 1	CCB83	-24	0.0152	-245.1		PPB	28 Mar 2019 20:59:47	U	SMPL
CRA - 1	CRA	190	0.0657	-245.1		PPB	28 Mar 2019 21:02:02	U	SMPL
HighStd - 1	HighStd	42060	9.9458	-245.1		PPB	28 Mar 2019 21:04:17	U	SMPL
ChkStd - 1	ChkStd	31521	7.4589	-245.1		PPB	28 Mar 2019 21:06:33	U	SMPL
PB118392BL - 1	PBW	-50	0.0091	-245.1		PPB	28 Mar 2019 21:08:49	U	SMPL
PB118392BS - 1	LCSW	17632	4.1815	-245.1		PPB	28 Mar 2019 21:11:05	U	SMPL
K1560-10 - 1	LOD-MDL-WATER-04-QT1-2019	189	0.0655	-245.1		PPB	28 Mar 2019 21:26:31	U	SMPL
K1560-11 - 1	LOQ-WATER-05-QT1-2019	1960	0.4834	-245.1		PPB	28 Mar 2019 21:28:47	U	SMPL
CCV84 - 1	CCV84	22059	5.2261	-245.1		PPB	28 Mar 2019 21:31:04	U	SMPL
CCB84 - 1	CCB84	-39	0.0117	-245.1		PPB	28 Mar 2019 21:33:21	U	SMPL