



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB101723

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	03/29/2019	09:35
2	CAL2	1	Water	NA	NA	20.2	7.01	03/29/2019	09:36
3	CAL3	1	Water	NA	NA	20.2	10.03	03/29/2019	09:37
4	ICV	1	Water	NA	NA	20.2	7.00	03/29/2019	09:39
5	CCV1	1	Water	NA	NA	20.2	2.01	03/29/2019	09:40
6	K1833-20	1	Water	NA	NA	23.0	7.99	03/29/2019	09:41
7	K2158-01	1	Water	NA	NA	16.6	5.32	03/29/2019	09:43
8	K2158-02	1	Water	NA	NA	17.3	5.32	03/29/2019	09:45
9	K2158-02DUP	1	Water	NA	NA	17.4	5.33	03/29/2019	09:46
10	CCV2	1	Water	NA	NA	20.1	12.02	03/29/2019	09:48

LB 101723

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph k2158

WorkList ID : 124132

Date : 03-29-2019 09:21:31

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/28/2019	Water	K2158-01 -B	pH	Cool 4 deg C	NYUL01	F22	SINK-4-ELGA-2WB	03/28/2019	SM 4500-H B
03/28/2019	Water	K2158-02 -B	pH	Cool 4 deg C	NYUL01	F22	SINK-6-ELGA-3EB	03/28/2019	SM 4500-H B

Date/Time 03-29-19 9:25
Received by: JO
Relinquished by: [Signature]

Date/Time 03-29-19 10:30
Received by: [Signature]
Relinquished by: JO