



Analytical Summary Report

Analysis Method: 9041A

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB101726

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467
0.0-3.0 pH paper refill	W1931
3.0-5.5 pH paper refill	W1934
Test Paper,PH Short Range 9.0/10.0	W1939
Test Paper, PH, Short Range 6.0/8.0	W2350
0.0-3.0 pH paper refill	W1931
10.0-12.0 pH paper refill	W1940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	23.0	4.0	03/29/2019	09:35
2	CAL2	1	Water	NA	NA	23.0	7.0	03/29/2019	09:36
3	CAL3	1	Water	NA	NA	23.0	10.0	03/29/2019	09:37
4	ICV	1	Water	NA	NA	23.0	7.0	03/29/2019	09:39
5	CCV1	1	Water	NA	NA	23.0	2.0	03/29/2019	09:40
6	K1833-22	1	Water	NA	NA	23.0	8.0	03/29/2019	09:41
7	CCV2	1	Water	NA	NA	23.0	12.0	03/29/2019	09:48