

Analytical Summary Report

Analysis Method: SM5540 C

ANALYST: AMANDEEP

Parameter: MBAS

SUPERVISOR REVIEW BY: apatel

Run Number: LB101735

Reagent/Standard	Lot/Loq #
MBAS CAL STD, 0.5PPM	WP73107
MBAS CCV STD, 1.0PPM	WP73115
MBAS ICV-LCS std 1.0PPM	WP73116
MBAS DAILY CHECK STD, 0.01PPM	WP73121
MBAS CAL STD 2 PPM	WP73114
MBAS CAL STD 1 PPM	WP73106
MBAS CAL STD 0.1 PPM	WP73110
mbas cal std 0.05 ppm	WP73111
MBAS CAL STD 0.01 PPM	WP73112
MBAS BLANK 0 PPM	WP73104
MBAS CAL STD 0.4 PPM	WP73109
MBAS CAL STD 0.2 PPM	WP73108
MBAS DAILY CHECK STD, 2.0PPM	WP73122
Chloroform, 4X4L	W2175

Intercept: 0.0008

Slope: 0.2234

Regression: 0.9998

Seq	Lab ID	True Value (mg/L)	Dilution	Reading (652 nm)	Anal Date	Anal Time
1	CAL1	0	1	0.000	03/29/2019	12:54
2	CAL2	0.01	1	0.002	03/29/2019	12:55
3	CAL3	0.05	1	0.011	03/29/2019	12:55
4	CAL4	0.1	1	0.027	03/29/2019	12:56
5	CAL5	0.2	1	0.048	03/29/2019	12:56
6	CAL6	0.4	1	0.090	03/29/2019	12:57
7	CAL7	0.5	1	0.112	03/29/2019	12:57
8	CAL8	1	1	0.219	03/29/2019	12:58
9	CAL9	2	1	0.435	03/29/2019	12:58



Analytical Summary Report

Analysis Method: SM5540 C

ANALYST: AMANDEEP

Parameter: MBAS

SUPERVISOR REVIEW BY: apatel

Run Number: LB101735

Seq	Lab ID	True Value (mg/L)	Dilution	Reading (652 nm)	Result (mg/L)	AnalDate	Anal Time
1	ICV	1	1	0.218	0.972	03/29/2019	12:59
2	ICB		1	0.000	-0.004	03/29/2019	13:00
3	CHK1	0.01	1	0.003	0.010	03/29/2019	13:01
4	CHK2	2	1	0.451	2.015	03/29/2019	13:02
5	CCV1	1	1	0.220	0.981	03/29/2019	13:02
6	CCB1		1	0.000	-0.004	03/29/2019	13:03
7	LB101735BL		1	0.000	-0.004	03/29/2019	13:04
8	LB101735BS	1	1	0.218	0.972	03/29/2019	13:04
9	K1833-30		1	0.171	0.762	03/29/2019	13:05
10	K1833-30DUP		1	0.172	0.766	03/29/2019	13:06
11	CCV2	1	1	0.219	0.977	03/29/2019	13:06
12	CCB2		1	0.000	-0.004	03/29/2019	13:07