


LB101755

7471B

INSTRUMENT ID: C01



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.033	0.033	46	0.000	46				
0.2	0.200	0.196	-0.004	1001	0.0 %	1001				
2.5	2.500	2.485	-0.015	14363	0.0 %	14363				
5.0	5.000	4.963	-0.037	28836	0.0 %	28836				
7.5	7.500	7.506	0.006	43678	0.0 %	43678				
10.0	10.000	10.018	0.018	58346	0.0 %	58346				

LB101755 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 29 Mar 2019 15:47:31

Sample ID	Extended ID	µ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	S0-0	46	-	0.0000	7471B	PPB	29 Mar 2019 15:47:47	S	Std
0.2 - 1	S02	1001	-	0.2000	7471B	PPB	29 Mar 2019 15:50:06	S	Std
2.5 - 1	S2-S	14363	-	2.5000	7471B	PPB	29 Mar 2019 15:52:24	S	Std
5.0 - 1	S5-Q	28836	-	5.0000	7471B	PPB	29 Mar 2019 15:54:41	S	Std
7.5 - 1	S7-S	43678	-	7.5000	7471B	PPB	29 Mar 2019 15:56:58	S	Std
10.0 - 1	S10-0	58346	-	10.0000	7471B	PPB	29 Mar 2019 16:11:41	S	Std
ICV78 - 1	ICV78	23036	3.9701	-	7471B	PPB	29 Mar 2019 16:15:23	U	SMPL
ICB78 - 1	ICB78	-46	0.0168	-	7471B	PPB	29 Mar 2019 16:17:38	U	SMPL
CCV92 - 1	CCV92	28538	4.9124	-	7471B	PPB	29 Mar 2019 16:19:54	U	SMPL
CCB92 - 1	CCB92	-61	0.0142	-	7471B	PPB	29 Mar 2019 16:22:09	U	SMPL
CRA - 1	CRA	1067	0.2074	-	7471B	PPB	29 Mar 2019 16:24:26	U	SMPL
HighStd - 1	HighStd	58720	10.0818	-	7471B	PPB	29 Mar 2019 16:26:42	U	SMPL
ChkStd - 1	ChkStd	42938	7.3788	-	7471B	PPB	29 Mar 2019 16:28:57	U	SMPL
PB118413BL - 1	PBS	-83	0.0104	-	7471B	PPB	29 Mar 2019 16:31:15	U	SMPL
PB118413BS - 1	LCSS	22043	3.8000	-	7471B	PPB	29 Mar 2019 16:36:59	U	SMPL
K1560-03 - 1	MDL-MDL-SOIL-03-QT1-2019	219	0.0622	-	7471B	PPB	29 Mar 2019 16:39:17	U	SMPL
PB118417BL - 1	PBS	0	0.0246	-	7471B	PPB	29 Mar 2019 16:41:34	U	SMPL
PB118417BS - 1	LCSS	21373	3.6853	-	7471B	PPB	29 Mar 2019 16:43:51	U	SMPL
K1685-09 - 1	AU-06-032819-A	1523	0.2855	-	7471B	PPB	29 Mar 2019 16:46:08	U	SMPL
K1685-09DUP - 1	AU-06-032819-ADUP	1474	0.2771	-	7471B	PPB	29 Mar 2019 16:48:23	U	SMPL
CCV93 - 1	CCV93	28088	4.8354	-	7471B	PPB	29 Mar 2019 16:50:39	U	SMPL
CCB93 - 1	CCB93	-92	0.0089	-	7471B	PPB	29 Mar 2019 16:52:55	U	SMPL
K1685-09MS - 1	AU-06-032819-AMS	25579	4.4056	-	7471B	PPB	29 Mar 2019 16:58:32	U	SMPL
K1685-09MSD - 1	AU-06-032819-AMSD	25154	4.3329	-	7471B	PPB	29 Mar 2019 17:00:46	U	SMPL
K1685-11 - 1	AU-06-032819-B	2480	0.4494	-	7471B	PPB	29 Mar 2019 17:03:03	U	SMPL
K1685-13 - 1	AU-06-032819-C	1528	0.2864	-	7471B	PPB	29 Mar 2019 17:05:19	U	SMPL
K1698-03 - 1	TR-05-032819	5605	0.9846	-	7471B	PPB	29 Mar 2019 17:07:35	U	SMPL
K2138-01 - 1	S6A(2-10)COMP	777	0.1577	-	7471B	PPB	29 Mar 2019 17:09:51	U	SMPL
K2138-03 - 1	S6A(12-20)COMP	154	0.0510	-	7471B	PPB	29 Mar 2019 17:12:07	U	SMPL
K2138-05 - 1	S6A/S6B(21-25)COMP	174	0.0545	-	7471B	PPB	29 Mar 2019 17:14:24	U	SMPL
K2138-07 - 1	S6B(2-10)COMP	441	0.1002	-	7471B	PPB	29 Mar 2019 17:16:41	U	SMPL
K2138-09 - 1	S6B(12-10)COMP	235	0.0649	-	7471B	PPB	29 Mar 2019 17:18:57	U	SMPL
CCV94 - 1	CCV94	28581	4.9198	-	7471B	PPB	29 Mar 2019 17:21:13	U	SMPL
CCB94 - 1	CCB94	-113	0.0053	-	7471B	PPB	29 Mar 2019 17:23:28	U	SMPL
K2139-01 - 1	S7B(2-10)COMP	848	0.1699	-	7471B	PPB	29 Mar 2019 17:25:45	U	SMPL
K2139-03 - 1	S7B(12-20)COMP	431	0.0985	-	7471B	PPB	29 Mar 2019 17:28:01	U	SMPL
K2153-21 - 1	SP-1	45585	7.8321	-	7471B	PPB	29 Mar 2019 17:30:17	U	SMPL
K2153-25 - 1	SP-2	12159	2.1072	-	7471B	PPB	29 Mar 2019 17:32:33	U	SMPL
K2155-02 - 1	SSSI-0328-S-DH-5-T	12875	2.2298	-	7471B	PPB	29 Mar 2019 17:34:49	U	SMPL
K2155-04 - 1	SSSI-0328-S-DH-5-B	17719	3.0594	-	7471B	PPB	29 Mar 2019 17:37:06	U	SMPL
K1685-09LX5 - 1		230	0.0640	-	7471B	PPB	29 Mar 2019 17:39:23	U	SMPL
K1685-09A - 1		25092	4.3222	-	7471B	PPB	29 Mar 2019 17:41:40	U	SMPL
CCV95 - 1	CCV95	28602	4.9234	-	7471B	PPB	29 Mar 2019 17:43:57	U	SMPL
CCB95 - 1	CCB95	-89	0.0094	-	7471B	PPB	29 Mar 2019 17:46:13	U	SMPL