



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB101759

Reagent/Standard	Lot/Log #
COD Digestion Vials Low Level 0-150Mg/L	W2469
COD ICV-LCS std, 50ppm	WP72786
COD calibration std. 100 ppm	WP72783
COD calibration std. 10 ppm	WP72781
COD calibration std. 150 ppm	WP72784
COD calibration std. 50 ppm	WP72782
COD calibration std. 0 ppm	WP72780
COD ICV-LCS std, 50ppm	WP73162
COD CCV std, 50ppm	WP73161

Temp In (C): 150	Date In: 03/29/2019	Time In: 12:00
Temp Out (C): 150	Date Out: 03/29/2019	Time Out: 14:00

Intercept: -0.1952

Slope: 1.0031

Regression: 0.9999

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	03/21/2019	12:40
2	CAL2	10	1	Water	10.000	03/21/2019	12:40
3	CAL3	50	1	Water	49.000	03/21/2019	12:41
4	CAL4	100	1	Water	101.000	03/21/2019	12:41
5	CAL5	150	1	Water	150.000	03/21/2019	12:42



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB101759

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50			1	Water	50.000	50.040	03/21/2019	12:42
2	ICB				1	Water	0.000	0.195	03/21/2019	12:43
3	CCV1	50			1	Water	51.000	51.037	03/29/2019	15:00
4	CCB1				1	Water	0.000	0.195	03/29/2019	15:00
5	LB101759BL				1	Water	0.000	0.195	03/29/2019	15:01
6	LB101759BS	50			1	Water	51.000	51.037	03/29/2019	15:01
7	K2093-17				1	Water	72.000	71.972	03/29/2019	15:02
8	K2093-20				100	Water	24.000	24.120	03/29/2019	15:02
9	K2117-01				1	Water	0.000	0.195	03/29/2019	15:03
10	K2117-01DUP				1	Water	0.000	0.195	03/29/2019	15:03
11	K2117-01MS	50			1	Water	50.000	50.040	03/29/2019	15:03
12	K2117-01MSD	50			1	Water	51.000	51.037	03/29/2019	15:04
13	CCV2	50			1	Water	52.000	52.034	03/29/2019	15:04
14	CCB2				1	Water	0.000	0.195	03/29/2019	15:03