



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 4/1/2019

OVENTEMP IN Celsius(°C): 108
Time IN: 15:25
In Date: 03/31/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:35
Out Date: 04/01/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-1
Thermometer ID: %SOLIDS-OVEN

QC:LB101766

Lab ID	Client Sample ID	Dish#	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish + Dry Sample Wt(g) (C)	% Solid
K2173-01	T-3	1	1.11	9.52	8.97	93.5
K2175-01	T-1	2	1.13	9.62	9.26	95.8
K2175-02	T-2	3	1.14	9.98	9.38	93.2
K2175-03	T-3	4	1.19	9.88	9.48	95.4
K2175-04	T-4	5	1.15	9.95	9.26	92.2
K2175-05	T-5	6	1.13	9.71	9.31	95.3
K2178-01	COMP-16	7	1.12	9.94	8.99	89.2
K2178-02	COMP-17	8	1.14	9.98	9.06	89.6
K2178-03	COMP-18	9	1.11	9.67	8.78	89.6
K2178-07	VOC-16	10	1.16	6.37	5.92	91.4
K2178-08	VOC-17	11	1.16	6.86	6.54	94.4
K2178-09	VOC-18	12	1.13	6.63	5.97	88.0
K2182-01	S8A(2-10)COMP	13	1.14	9.64	9.19	94.7
K2182-03	S8A-S8B(12-16)COMP	14	1.14	9.91	9.45	94.8
K2182-05	S8B(2-10)COMP	15	1.18	9.51	9.09	95.0
K2182-07	S7C(2-10)COMP	16	1.18	9.74	8.96	90.9
K2182-09	S7C(12-20)COMP	17	1.13	9.76	9.34	95.1
K2182-11	S8A(10)	18	1.16	9.74	9.34	95.3
K2182-13	S8A(12)	19	1.16	9.88	9.51	95.8
K2182-15	S8B(10)	20	1.16	9.57	9.19	95.5
K2182-17	S7C(10)	21	1.14	9.6	9.16	94.8
K2182-19	S7C(16)	22	1.18	9.81	9.1	91.8
K2183-01	FK-0D015-03-002	23	1.18	9.92	9.47	94.9
K2183-02	FK-0D015-03-002	24	1.17	9.75	9.39	95.8
K2184-01	FK-0D015-03-002	25	1.18	9.97	9.38	93.3
K2184-02	FK-0D015-03-002	26	1.18	9.97	9.38	93.3
K2185-01	P-4-OW(0-10)	27	1.13	9.71	8.37	84.4
K2185-02	P-4-OW(0-10)	28	1.15	9.96	8.53	83.8
K2186-01	TP-TRACK-89-90-SOIL	29	1.19	9.72	8.5	85.7
K2186-03	TP-TRACK-89-90-SOIL	30	1.14	9.65	8.51	86.6
K2186-05	TP-TRACK-86-87-SOIL	31	1.16	9.89	9.16	91.6
K2186-07	TP-TRACK-86-87-SOIL	32	1.17	9.75	8.79	88.8
K2186-09	BELT-PARKWAY-STAGING-YA	33	1.18	9.61	9.02	93.0
K2186-11	BELT-PARKWAY-STAGING-YA	34	1.14	9.98	9.37	93.1
K2186-13	BELT-PARKWAY-STAGING-YA	35	1.11	9.54	9.09	94.7
K2186-15	BELT-PARKWAY-STAGING-YA	36	1.15	9.95	9.57	95.7
K2186-17	BELT-PARKWAY-STAGING-YA	37	1.16	9.76	9.47	96.6
K2186-19	BELT-PARKWAY-STAGING-YA	38	1.11	9.8	9.56	97.2
K2186-21	TP-TRACK-78-79	39	1.13	9.94	8.22	80.5



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 4/1/2019

OVENTEMP IN Celsius(°C): 108
Time IN: 15:25
In Date: 03/31/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:35
Out Date: 04/01/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-1
Thermometer ID: %SOLIDS-OVEN

QC:LB101766

<u>Lab ID</u>	<u>Client Sample ID</u>	<u>Dish#</u>	<u>Dish Wt(g)</u> <u>(A)</u>	<u>Dish +</u> <u>Sample Wt(g)</u> <u>(B)</u>	<u>Dish + Dry</u> <u>Sample Wt(g)</u> <u>(C)</u>	<u>% Solid</u>
K2186-23	TP-TRACK-78-79	40	1.13	9.71	8.23	82.8

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$