

LB101783


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

43401



A= 0.0000e+000

B= 2.2992e-004

C= 6.7355e-003

Rho= 0.9997918

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.024	0.024	76	0.000	76				
0.2	0.200	0.236	0.036	998	0.0 %	998				
2.5	2.500	2.479	-0.021	10753	0.0 %	10753				
5.0	5.000	4.863	-0.137	21120	0.0 %	21120				
7.5	7.500	7.612	0.112	33079	0.0 %	33079				
10.0	10.000	9.986	-0.014	43401	0.0 %	43401				

LB101783 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 01 Apr 2019 13:38:53

Sample ID	Extended ID	µ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50-0	76	-	0.00007470A	-7470A	PPB	01 Apr 2019 13:43:40	S	Std
0.2 - 1	50-2	998	-	0.20007470A	-7470A	PPB	01 Apr 2019 13:45:48	S	Std
2.5 - 1	52-5	10753	-	2.50007470A	-7470A	PPB	01 Apr 2019 13:47:56	S	Std
5.0 - 1	55-0	21120	-	5.00007470A	-7470A	PPB	01 Apr 2019 13:50:04	S	Std
7.5 - 1	57-5	33079	-	7.50007470A	-7470A	PPB	01 Apr 2019 13:52:11	S	Std
10.0 - 1	510-0	43401	-	10.00007470A	-7470A	PPB	01 Apr 2019 13:54:18	S	Std
ICV80 - 1	ICV80	17333	3.9920	-7470A	-7470A	PPB	01 Apr 2019 13:57:48	U	SMPL
ICB80 - 1	ICB80	-40	-0.0025	-7470A	-7470A	PPB	01 Apr 2019 13:59:53	U	SMPL
CCV98 - 1	CCV98	21133	4.8657	-7470A	-7470A	PPB	01 Apr 2019 14:01:59	U	SMPL
CCB98 - 1	CCB98	-6	0.0054	-7470A	-7470A	PPB	01 Apr 2019 14:04:05	U	SMPL
CRA - 1	CRA	936	0.2219	-7470A	-7470A	PPB	01 Apr 2019 14:06:11	U	SMPL
HighStd - 1	HighStd	43426	9.9913	-7470A	-7470A	PPB	01 Apr 2019 14:08:17	U	SMPL
ChkStd - 1	ChkStd	30065	6.9193	-7470A	-7470A	PPB	01 Apr 2019 14:10:23	U	SMPL
PB118481BL - 1	PBW	-67	-0.0087	-7470A	-7470A	PPB	01 Apr 2019 14:12:30	U	SMPL
PB118481BS - 1	LCSW	16581	3.8191	-7470A	-7470A	PPB	01 Apr 2019 14:14:37	U	SMPL
K2056-05 - 1	CS-7	-35	-0.0013	-7470A	-7470A	PPB	01 Apr 2019 14:16:44	U	SMPL
K2056-05DUP - 1	CS-7DUP	18	0.0109	-7470A	-7470A	PPB	01 Apr 2019 14:18:51	U	SMPL
K2056-05MS - 1	CS-7MS	17585	4.0499	-7470A	-7470A	PPB	01 Apr 2019 14:20:59	U	SMPL
K2056-05MSD - 1	CS-7MSD	17488	4.0276	-7470A	-7470A	PPB	01 Apr 2019 14:23:07	U	SMPL
K2056-06 - 1	CS-8	-16	0.0031	-7470A	-7470A	PPB	01 Apr 2019 14:25:13	U	SMPL
CCV99 - 1	CCV99	21435	4.9351	-7470A	-7470A	PPB	01 Apr 2019 14:27:19	U	SMPL
CCB99 - 1	CCB99	-15	0.0033	-7470A	-7470A	PPB	01 Apr 2019 14:29:26	U	SMPL
K2056-07 - 1	CS-9	23	0.0120	-7470A	-7470A	PPB	01 Apr 2019 14:31:32	U	SMPL
K2056-08 - 1	CS-10	35	0.0148	-7470A	-7470A	PPB	01 Apr 2019 14:33:38	U	SMPL
K2170-01 - 1	OIL-WATER-SEPERATOR-2	0	0.0067	-7470A	-7470A	PPB	01 Apr 2019 14:35:45	U	SMPL
K2178-04 - 1	COMP-16	81	0.0254	-7470A	-7470A	PPB	01 Apr 2019 14:37:51	U	SMPL
K2178-05 - 1	COMP-17	25	0.0125	-7470A	-7470A	PPB	01 Apr 2019 14:39:58	U	SMPL
K2178-06 - 1	COMP-18	53	0.0189	-7470A	-7470A	PPB	01 Apr 2019 14:42:05	U	SMPL
K2182-02 - 1	S8A(2-10)COMP	17	0.0106	-7470A	-7470A	PPB	01 Apr 2019 14:44:12	U	SMPL
K2182-04 - 1	S8A-S8B(12-16)COMP	30	0.0136	-7470A	-7470A	PPB	01 Apr 2019 14:46:20	U	SMPL
K2182-06 - 1	S8B(2-10)COMP	26	0.0127	-7470A	-7470A	PPB	01 Apr 2019 14:48:28	U	SMPL
K2182-08 - 1	S7C(2-10)COMP	13	0.0097	-7470A	-7470A	PPB	01 Apr 2019 14:50:35	U	SMPL
CCV100 - 1	CCV100	21453	4.9392	-7470A	-7470A	PPB	01 Apr 2019 14:52:41	U	SMPL
CCB100 - 1	CCB100	-1	0.0065	-7470A	-7470A	PPB	01 Apr 2019 14:54:48	U	SMPL
K2182-10 - 1	S7C(12-20)COMP	21	0.0116	-7470A	-7470A	PPB	01 Apr 2019 14:56:54	U	SMPL
K2186-02 - 1	TP-TRACK-89-90-SOIL	65	0.0217	-7470A	-7470A	PPB	01 Apr 2019 14:59:01	U	SMPL
K2186-06 - 1	TP-TRACK-86-87-SOIL	42	0.0164	-7470A	-7470A	PPB	01 Apr 2019 15:01:07	U	SMPL
K2186-10 - 1	BELT-PARKWAY-STAGING-YARD-1	17	0.0106	-7470A	-7470A	PPB	01 Apr 2019 15:03:14	U	SMPL
K2186-14 - 1	BELT-PARKWAY-STAGING-YARD-2	2	0.0072	-7470A	-7470A	PPB	01 Apr 2019 15:05:20	U	SMPL
K2186-18 - 1	BELT-PARKWAY-STAGING-YARD-3	31	0.0139	-7470A	-7470A	PPB	01 Apr 2019 15:07:28	U	SMPL
K2186-22 - 1	TP-TRACK-78-79	28	0.0132	-7470A	-7470A	PPB	01 Apr 2019 15:09:35	U	SMPL
PB118482BL - 1	PBW	61	0.0208	-7470A	-7470A	PPB	01 Apr 2019 15:11:43	U	SMPL
PB118482BS - 1	LCSW	15959	3.6760	-7470A	-7470A	PPB	01 Apr 2019 15:13:51	U	SMPL
K1689-04 - 1	NB-08-032919-A	245	0.0631	-7470A	-7470A	PPB	01 Apr 2019 15:15:57	U	SMPL
CCV101 - 1	CCV101	23174	5.3349	-7470A	-7470A	PPB	01 Apr 2019 15:18:04	U	SMPL
CCB101 - 1	CCB101	-16	0.0031	-7470A	-7470A	PPB	01 Apr 2019 15:20:11	U	SMPL
K1689-04DUP - 1	NB-08-032919-ADUP	298	0.0753	-7470A	-7470A	PPB	01 Apr 2019 15:22:17	U	SMPL
K1689-04MS - 1	NB-08-032919-AMS	18152	4.1803	-7470A	-7470A	PPB	01 Apr 2019 15:24:24	U	SMPL
K1689-04MSD - 1	NB-08-032919-AMSD	18249	4.2026	-7470A	-7470A	PPB	01 Apr 2019 15:26:31	U	SMPL
K1689-06 - 1	NB-08-032919-B	254	0.0651	-7470A	-7470A	PPB	01 Apr 2019 15:28:37	U	SMPL
K1692-06 - 1	OK-03-032919	275	0.0700	-7470A	-7470A	PPB	01 Apr 2019 15:30:44	U	SMPL
K2171-02 - 1	GW-COMP	306	0.0771	-7470A	-7470A	PPB	01 Apr 2019 15:32:51	U	SMPL
K2174-02 - 1	RBR200030	309	0.0778	-7470A	-7470A	PPB	01 Apr 2019 15:34:58	U	SMPL
K2179-04 - 1	TP-4S	306	0.0771	-7470A	-7470A	PPB	01 Apr 2019 15:37:06	U	SMPL
K2179-08 - 1	TP-4D	309	0.0778	-7470A	-7470A	PPB	01 Apr 2019 15:39:14	U	SMPL
K2179-12 - 1	TP-8S	292	0.0739	-7470A	-7470A	PPB	01 Apr 2019 15:41:21	U	SMPL
CCV102 - 1	CCV102	23299	5.3637	-7470A	-7470A	PPB	01 Apr 2019 15:43:27	U	SMPL
CCB102 - 1	CCB102	-39	-0.0022	-7470A	-7470A	PPB	01 Apr 2019 15:45:34	U	SMPL
K2179-16 - 1	TP-8D	294	0.0743	-7470A	-7470A	PPB	01 Apr 2019 15:47:40	U	SMPL
K2179-20 - 1	TP-5S	284	0.0720	-7470A	-7470A	PPB	01 Apr 2019 15:49:47	U	SMPL
K2180-04 - 1	TP-7	309	0.0778	-7470A	-7470A	PPB	01 Apr 2019 15:51:53	U	SMPL
K2180-08 - 1	TP-6S	289	0.0732	-7470A	-7470A	PPB	01 Apr 2019 15:54:00	U	SMPL
K2180-12 - 1	TP-6D	280	0.0711	-7470A	-7470A	PPB	01 Apr 2019 15:56:07	U	SMPL
K2180-16 - 1	TP-9	277	0.0704	-7470A	-7470A	PPB	01 Apr 2019 15:58:14	U	SMPL
K2180-20 - 1	TP-10	290	0.0734	-7470A	-7470A	PPB	01 Apr 2019 16:00:21	U	SMPL
K2180-24 - 1	TP-10D	274	0.0697	-7470A	-7470A	PPB	01 Apr 2019 16:02:29	U	SMPL
K2180-28 - 1	TP-1	276	0.0702	-7470A	-7470A	PPB	01 Apr 2019 16:04:37	U	SMPL
K2180-32 - 1	TP-3	291	0.0736	-7470A	-7470A	PPB	01 Apr 2019 16:06:43	U	SMPL
CCV103 - 1	CCV103	23514	5.4131	-7470A	-7470A	PPB	01 Apr 2019 16:08:51	U	SMPL
CCB103 - 1	CCB103	-18	0.0026	-7470A	-7470A	PPB	01 Apr 2019 16:10:57	U	SMPL
PB118483BL - 1	PBW	285	0.0723	-7470A	-7470A	PPB	01 Apr 2019 16:13:05	U	SMPL
PB118483BS - 1	LCSW	16454	3.7899	-7470A	-7470A	PPB	01 Apr 2019 16:15:12	U	SMPL
K1685-10 - 1	AU-06-032819-A	295	0.0746	-7470A	-7470A	PPB	01 Apr 2019 16:17:19	U	SMPL
K1685-10DUP - 1	AU-06-032819-ADUP	343	0.0856	-7470A	-7470A	PPB	01 Apr 2019 16:19:26	U	SMPL
K1685-10MS - 1	AU-06-032819-AMS	18597	4.2826	-7470A	-7470A	PPB	01 Apr 2019 16:21:33	U	SMPL
K1685-10MSD - 1	AU-06-032819-AMSD	18488	4.2575	-7470A	-7470A	PPB	01 Apr 2019 16:23:40	U	SMPL
K1685-12 - 1	AU-06-032819-B	76	0.0242	-7470A	-7470A	PPB	01 Apr 2019 16:25:48	U	SMPL
K1685-14 - 1	AU-06-032819-C	581	0.1403	-7470A	-7470A	PPB	01 Apr 2019 16:27:56	U	SMPL
K1697-14 - 1	SU-02-032719-A	502	0.1222	-7470A	-7470A	PPB	01 Apr 2019 16:30:04	U	SMPL
K1697-16 - 1	SU-02-032719-B	437	0.1072	-7470A	-7470A	PPB	01 Apr 2019 16:32:10	U	SMPL
CCV104 - 1	CCV104	20113	4.6311	-7470A	-7470A	PPB	01 Apr 2019 16:34:18	U	SMPL
CCB104 - 1	CCB104	-13	0.0037	-7470A	-7470A	PPB	01 Apr 2019 16:36:25	U	SMPL
K1698-04 - 1	TR-05-032819	375	0.0930	-7470A	-7470A	PPB	01 Apr 2019 16:38:32	U	SMPL
K2125-04 - 1	SOIL-COMP	340	0.0849	-7470A	-7470A	PPB	01 Apr 2019 16:40:39	U	SMPL
K2138-02 - 1	S6A(2-10)COMP	14	0.0100	-7470A	-7470A	PPB	01 Apr 2019 16:42:46	U	SMPL
K2138-04 - 1	S6A(12-20)COMP	519	0.1261	-7470A	-7470A	PPB	01 Apr 2019 16:44:54	U	SMPL
K2138-06 - 1	S6A/S6B(21-25)COMP	516	0.1254	-7470A	-7470A	PPB	01 Apr 2019 16:47:01	U	SMPL
K2138-08 - 1	S6B(2-10)COMP	426	0.1047	-7470A	-7470A	PPB	01 Apr 2019 16:49:08	U	SMPL
K2138-10 - 1	S6B(12-20)COMP	362	0.0900	-7470A	-7470A	PPB	01 Apr 2019 16:51:15	U	SMPL
K2139-02 - 1	S7B(2-10)COMP	320	0.0803	-7470A	-7470A	PPB	01 Apr 2019 16:53:23	U	SMPL
K2139-04 - 1	S7B(12-20)COMP	304	0.0766	-7470A	-7470A	PPB	01 Apr 2019 16:55:31	U	SMPL
K2153-24 - 1	SP-1	273	0.0695	-7470A	-7470A	PPB	01 Apr 2019 16:57:38	U	SMPL
CCV105 - 1	CCV105	21844	5.0291	-7470A	-7470A	PPB	01 Apr 2019 16:59:45	U	SMPL

LB101783 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 01 Apr 2019 13:38:53

Sample ID	Extended ID	μ Abs.	Conc.	Stnd Conc	Method	Units	Date	Type	Type
CCB105 - 1	CCB105	-12	0.0040	-7470A	PPB		01 Apr 2019 17:01:52	U	SMPL
K2153-28 - 1	SP-2	248	0.0638	-7470A	PPB		01 Apr 2019 17:03:59	U	SMPL
PB118481TB - 1	PB118452TB	285	0.0723	-7470A	PPB		01 Apr 2019 17:06:06	U	SMPL
PB118482TB - 1	PB118444TB	317	0.0796	-7470A	PPB		01 Apr 2019 17:08:13	U	SMPL
PB118483TB - 1	PB118369TB	307	0.0773	-7470A	PPB		01 Apr 2019 17:10:20	U	SMPL
K2056-05LX5 - 1		-9	0.0047	-7470A	PPB		01 Apr 2019 17:12:27	U	SMPL
K2056-05A - 1		18104	4.1692	-7470A	PPB		01 Apr 2019 17:14:34	U	SMPL
K1689-04LX5 - 1		4	0.0077	-7470A	PPB		01 Apr 2019 17:16:41	U	SMPL
K1689-04A - 1		17909	4.1244	-7470A	PPB		01 Apr 2019 17:18:49	U	SMPL
K1685-10LX5 - 1		33	0.0143	-7470A	PPB		01 Apr 2019 17:20:57	U	SMPL
K1685-10A - 1		18486	4.2571	-7470A	PPB		01 Apr 2019 17:23:05	U	SMPL
CCV106 - 1	CCV106	21906	5.0434	-7470A	PPB		01 Apr 2019 17:25:12	U	SMPL
CCB106 - 1	CCB106	-70	-0.0094	-7470A	PPB		01 Apr 2019 17:27:19	U	SMPL