



Analytical Summary Report

Analysis Method: 9040C

Parameter: Corrosivity

Run Number: LB101792

Analyst By : JIGNESH

Supervisor Review By : apatel

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	04/01/2019	17:00
2	CAL2	1	Water	NA	NA	20.1	7.01	04/01/2019	17:01
3	CAL3	1	Water	NA	NA	20.1	10.02	04/01/2019	17:03
4	ICV	1	Water	NA	NA	20.1	7.01	04/01/2019	17:05
5	CCV1	1	Water	NA	NA	20.2	2.02	04/01/2019	17:06
6	K1695-02	1	Water	NA	NA	17.5	9.00	04/01/2019	17:07
7	K1695-02DUP	1	Water	NA	NA	17.6	9.01	04/01/2019	17:08
8	CCV2	1	Water	NA	NA	20.1	12.01	04/01/2019	17:15

28101892

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY W K1695

WorkList ID : 124235

Date : 04-01-2019 16:46:33

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/01/2019	Water	K1695-02	Corrosivity	Cool 4 deg C	PSEG01	K31	SV-28399L	04/01/2019	9040C

Date/Time 04-01-19 16:50
 Received by: Jg
 Relinquished by: CB

Date/Time 04-01-19 17:30
 Received by: CB
 Relinquished by: Jg