



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 4/2/2019

OVENTEMP IN Celsius(°C): 108
Time IN: 17:20
In Date: 04/01/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:11
Out Date: 04/02/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-1
Thermometer ID: %SOLIDS-OVEN

QC:LB101794

Lab ID	Client Sample ID	Dish#	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish + Dry Sample Wt(g) (C)	% Solid
K1695-04	DPC-91723-SO-TPS5-A	1	1.13	9.65	8.88	91.0
K1695-05	DPC-91723-SO-TPS5-A-1	2	1.13	9.65	8.88	91.0
K1695-06	DPC-91723-SO-TPS5-A-2	3	1.13	9.65	8.88	91.0
K1695-07	DPC-91723-SO-TPS5-A	4	1.13	9.65	8.88	91.0
K1695-08	DPC-91723-SO-TPS5-B	5	1.16	9.74	9.00	91.4
K1695-09	DPC-91723-SO-TPS5-B-1	6	1.16	9.74	9.00	91.4
K1695-10	DPC-91723-SO-TPS5-B-2	7	1.16	9.74	9.00	91.4
K1695-11	DPC-91723-SO-TPS5-B	8	1.16	9.74	9.00	91.4
K2112-04	DR-WC-001	9	1.15	9.97	6.74	63.4
K2190-01	VW-SP-1	10	1.15	9.95	9.29	92.5
K2190-02	VW-SP-1	11	1.15	9.95	9.29	92.5
K2190-04	VW-SP-2	12	1.13	9.74	8.67	87.6
K2190-05	VW-SP-2	13	1.13	9.74	8.67	87.6
K2196-01	HD-01-040119	14	1.18	9.85	9.06	90.9
K2205-01	CONCRETE-A	15	1.00	2.00	2.00	100.0
K2205-02	CONCRETE-B	16	1.00	2.00	2.00	100.0
K2209-01	E-6	17	1.16	9.95	8.01	77.9
K2209-02	E-7	18	1.14	9.65	8.48	86.3
K2209-03	E-5	19	1.19	9.83	8.46	84.1
K2209-04	E-3	20	1.14	9.65	8.24	83.4
K2209-05	E-3A	21	1.1	9.81	8.4	83.8
K2209-06	E-8	22	1.17	9.88	9.06	90.6
K2209-07	E-9	23	1.18	9.97	8.48	83.0
K2209-08	E-4	24	1.14	9.94	8.09	79.0
K2209-09	E-1	25	1.13	9.96	8.42	82.6
K2209-10	E-2	26	1.17	9.89	7.71	75.0

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$