

LB101801


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

44863



A= 0.0000e+000

B= 2.2466e-004

C= -3.5408e-002

Rho= 0.9999163

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.038	0.038	328	0.000	328				
0.2	0.200	0.245	0.045	1248	0.0 %	1248				
2.5	2.500	2.431	-0.069	10980	0.0 %	10980				
5.0	5.000	4.940	-0.060	22147	0.0 %	22147				
7.5	7.500	7.502	0.002	33551	0.0 %	33551				
10.0	10.000	10.043	0.043	44863	0.0 %	44863				

LB101801 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 02 Apr 2019 10:28:51

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50-0	328	-	0.0000	7470A	PPB	02 Apr 2019 10:41:22	S	Std
0.2 - 1	50-2	1248	-	0.2000	7470A	PPB	02 Apr 2019 10:43:30	S	Std
2.5 - 1	52-5	10980	-	2.5000	7470A	PPB	02 Apr 2019 10:45:38	S	Std
5.0 - 1	55-0	22147	-	5.0000	7470A	PPB	02 Apr 2019 10:47:46	S	Std
7.5 - 1	57-5	33551	-	7.5000	7470A	PPB	02 Apr 2019 10:49:53	S	Std
10.0 - 1	510-0	44863	-	10.0000	7470A	PPB	02 Apr 2019 10:52:00	S	Std
ICV81 - 1	ICV81	18273	4.0697	-7470A	PPB	PPB	02 Apr 2019 10:55:53	U	SMPL
ICB81 - 1	ICB81	263	0.0237	-7470A	PPB	PPB	02 Apr 2019 11:01:06	U	SMPL
CCV08 - 1	CCV08	21842	4.8715	-7470A	PPB	PPB	02 Apr 2019 11:03:11	U	SMPL
CCB08 - 1	CCB08	29	-0.0289	-7470A	PPB	PPB	02 Apr 2019 11:05:17	U	SMPL
CRA - 1	CRA	1155	0.2241	-7470A	PPB	PPB	02 Apr 2019 11:07:23	U	SMPL
HighStd - 1	HighStd	44026	9.8553	-7470A	PPB	PPB	02 Apr 2019 11:09:29	U	SMPL
ChkStd - 1	ChkStd	30331	6.7786	-7470A	PPB	PPB	02 Apr 2019 11:11:35	U	SMPL
PB118529BL - 1	PBW	-36	-0.0435	-7470A	PPB	PPB	02 Apr 2019 11:13:42	U	SMPL
PB118529BS - 1	LCSW	16591	3.6919	-7470A	PPB	PPB	02 Apr 2019 11:15:49	U	SMPL
K1695-01 - 1	SV-28399L	154	-0.0008	-7470A	PPB	PPB	02 Apr 2019 11:17:56	U	SMPL
K1695-01DUP - 1	SV-28399LDUP	184	0.0059	-7470A	PPB	PPB	02 Apr 2019 11:20:04	U	SMPL
K1695-01MS - 1	SV-28399LMS	17726	3.9468	-7470A	PPB	PPB	02 Apr 2019 11:22:11	U	SMPL
K1695-01MSD - 1	SV-28399LMSD	18028	4.0147	-7470A	PPB	PPB	02 Apr 2019 11:24:19	U	SMPL
K1696-01 - 1	COMP-1	837	0.1526	-7470A	PPB	PPB	02 Apr 2019 11:26:26	U	SMPL
CCV09 - 1	CCV09	21791	4.8601	-7470A	PPB	PPB	02 Apr 2019 11:28:31	U	SMPL
CCB09 - 1	CCB09	46	-0.0251	-7470A	PPB	PPB	02 Apr 2019 11:30:37	U	SMPL
K1696-02 - 1	COMP-2	231	0.0165	-7470A	PPB	PPB	02 Apr 2019 11:32:44	U	SMPL
K1696-03 - 1	21860	1409	0.2811	-7470A	PPB	PPB	02 Apr 2019 11:34:50	U	SMPL
K2207-01 - 1	MH-6	60	-0.0219	-7470A	PPB	PPB	02 Apr 2019 11:36:56	U	SMPL
K2208-02 - 1	EFFL-2	806	0.1457	-7470A	PPB	PPB	02 Apr 2019 11:39:04	U	SMPL
K1695-01LX5 - 1		-6	-0.0368	-7470A	PPB	PPB	02 Apr 2019 11:41:11	U	SMPL
K1695-01A - 1		17784	3.9599	-7470A	PPB	PPB	02 Apr 2019 11:43:18	U	SMPL
CCV10 - 1	CCV10	21612	4.8199	-7470A	PPB	PPB	02 Apr 2019 11:45:25	U	SMPL
CCB10 - 1	CCB10	-20	-0.0399	-7470A	PPB	PPB	02 Apr 2019 11:47:33	U	SMPL