

LB101804


245.1

INSTRUMENT ID : CV1

Linear

 μ Abs.:

41785



A= 0.0000e+000

B= 2.3897e-004

C= -1.5853e-001

Rho= 0.9996517

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.046	-0.046	469	0.000	469				
0.05	0.050	0.037	-0.013	819	0.0 %	819				
0.2	0.200	0.154	-0.046	1307	0.0 %	1307				
2.5	2.500	2.545	0.045	11314	0.0 %	11314				
5.0	5.000	5.076	0.076	21904	0.0 %	21904				
7.5	7.500	7.658	0.158	32709	0.0 %	32709				
10.0	10.000	9.827	-0.173	41785	0.0 %	41785				

LB101804 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 02 Apr 2019 11:59:26

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	469	-	0.0000	245.1	PPB	02 Apr 2019 12:01:12	S	Std
0.05 - 1	50.05	819	-	0.0500	245.1	PPB	02 Apr 2019 12:03:29	S	Std
0.2 - 1	50.2	1307	-	0.2000	245.1	PPB	02 Apr 2019 12:05:47	S	Std
2.5 - 1	52.5	11314	-	2.5000	245.1	PPB	02 Apr 2019 12:08:04	S	Std
5.0 - 1	55.0	21904	-	5.0000	245.1	PPB	02 Apr 2019 12:10:21	S	Std
7.5 - 1	57.5	32709	-	7.5000	245.1	PPB	02 Apr 2019 12:12:37	S	Std
10.0 - 1	510.0	41785	-	10.0000	245.1	PPB	02 Apr 2019 12:26:01	S	Std
ICV82 - 1	ICV82	16161	3.7034	-245.1		PPB	02 Apr 2019 12:29:53	U	SMPL
ICB82 - 1	ICB82	396	-0.0639	-245.1		PPB	02 Apr 2019 12:32:10	U	SMPL
CCV11 - 1	CCV11	22081	5.1181	-245.1		PPB	02 Apr 2019 12:34:25	U	SMPL
CCB11 - 1	CCB11	418	-0.0586	-245.1		PPB	02 Apr 2019 12:36:40	U	SMPL
CRA - 1	CRA	934	0.0647	-245.1		PPB	02 Apr 2019 13:02:40	U	SMPL
HighStd - 1	HighStd	43174	10.1586	-245.1		PPB	02 Apr 2019 13:04:55	U	SMPL
ChkStd - 1	ChkStd	28676	6.6941	-245.1		PPB	02 Apr 2019 13:07:10	U	SMPL
PB118480BL - 1	PBW	374	-0.0692	-245.1		PPB	02 Apr 2019 13:09:27	U	SMPL
PB118480BS - 1	LCSW	17820	4.0999	-245.1		PPB	02 Apr 2019 13:11:44	U	SMPL
K2177-01 - 1	032819-LIFT-E-D-B	280	-0.0916	-245.1		PPB	02 Apr 2019 13:14:00	U	SMPL
K2177-01DUP - 1	032819-LIFT-E-D-BDUP	319	-0.0823	-245.1		PPB	02 Apr 2019 13:16:17	U	SMPL
K2177-01MS - 1	032819-LIFT-E-D-BMS	19008	4.3838	-245.1		PPB	02 Apr 2019 13:18:34	U	SMPL
K2177-01MSD - 1	032819-LIFT-E-D-BMSD	19110	4.4081	-245.1		PPB	02 Apr 2019 13:20:52	U	SMPL
K2177-02 - 1	032819-LIFT-E-D-M	278	-0.0921	-245.1		PPB	02 Apr 2019 13:23:07	U	SMPL
CCV12 - 1	CCV12	21373	4.9489	-245.1		PPB	02 Apr 2019 13:25:23	U	SMPL
CCB12 - 1	CCB12	399	-0.0632	-245.1		PPB	02 Apr 2019 13:27:38	U	SMPL
K2177-03 - 1	032819-LIFT-E-D-S	325	-0.0809	-245.1		PPB	02 Apr 2019 13:29:54	U	SMPL
K2177-04 - 1	032819-LIFT-E-U-B	348	-0.0754	-245.1		PPB	02 Apr 2019 13:32:10	U	SMPL
K2177-05 - 1	032819-LIFT-E-U-M	329	-0.0799	-245.1		PPB	02 Apr 2019 13:34:25	U	SMPL
K2177-06 - 1	032819-LIFT-E-U-S	307	-0.0852	-245.1		PPB	02 Apr 2019 13:36:41	U	SMPL
K2177-01LX5 - 1		46	-0.1475	-245.1		PPB	02 Apr 2019 13:38:57	U	SMPL
K2177-01A - 1		19273	4.4471	-245.1		PPB	02 Apr 2019 13:41:14	U	SMPL
CCV13 - 1	CCV13	21334	4.9396	-245.1		PPB	02 Apr 2019 13:43:30	U	SMPL
CCB13 - 1	CCB13	296	-0.0878	-245.1		PPB	02 Apr 2019 13:48:43	U	SMPL