

LB101812


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

61618



A= 0.0000e+000

B= 1.6232e-004

C= -1.1421e-001

Rho= 0.9994897

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.044	-0.044	434	0.000	434				
0.2	0.200	0.160	-0.040	1688	0.0 %	1688				
2.5	2.500	2.457	-0.043	15842	0.0 %	15842				
5.0	5.000	5.255	0.255	33079	0.0 %	33079				
7.5	7.500	7.484	-0.016	46809	0.0 %	46809				
10.0	10.000	9.888	-0.112	61618	0.0 %	61618				

LB101812 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 02 Apr 2019 15:31:16

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50-0	434	-	0.00007471B	PPB		02 Apr 2019 15:35:48	S	Std
0.2 - 1	50-2	1688	-	0.20007471B	PPB		02 Apr 2019 15:38:05	S	Std
2.5 - 1	52-5	15842	-	2.50007471B	PPB		02 Apr 2019 15:40:22	S	Std
5.0 - 1	55-0	33079	-	5.00007471B	PPB		02 Apr 2019 15:42:40	S	Std
7.5 - 1	57-5	46809	-	7.50007471B	PPB		02 Apr 2019 15:44:56	S	Std
10.0 - 1	510-0	61618	-	10.00007471B	PPB		02 Apr 2019 15:47:13	S	Std
ICV83 - 1	ICV83	25235	3.9819	-7471B	PPB		02 Apr 2019 15:50:36	U	SMPL
ICB83 - 1	ICB83	309	-0.0641	-7471B	PPB		02 Apr 2019 15:52:51	U	SMPL
CCV14 - 1	CCV14	32247	5.1201	-7471B	PPB		02 Apr 2019 15:55:07	U	SMPL
CCB14 - 1	CCB14	123	-0.0942	-7471B	PPB		02 Apr 2019 15:57:22	U	SMPL
CRA - 1	CRA	1632	0.1507	-7471B	PPB		02 Apr 2019 15:59:38	U	SMPL
HighStd - 1	HighStd	60866	9.7656	-7471B	PPB		02 Apr 2019 16:01:53	U	SMPL
ChkStd - 1	ChkStd	42752	6.8253	-7471B	PPB		02 Apr 2019 16:04:09	U	SMPL
PB118506BL - 1	PBS	328	-0.0610	-7471B	PPB		02 Apr 2019 16:09:13	U	SMPL
PB118506BS - 1	LCSS	21661	3.4018	-7471B	PPB		02 Apr 2019 16:11:29	U	SMPL
K2179-03 - 1	TP-4S	770	0.0108	-7471B	PPB		02 Apr 2019 16:13:45	U	SMPL
K2179-03DUP - 1	TP-4SDUP	768	0.0104	-7471B	PPB		02 Apr 2019 16:16:02	U	SMPL
K2179-03MS - 1	TP-4SMS	22952	3.6114	-7471B	PPB		02 Apr 2019 16:18:19	U	SMPL
K2179-03MSD - 1	TP-4MSD	23155	3.6443	-7471B	PPB		02 Apr 2019 16:20:36	U	SMPL
K2179-07 - 1	TP-4D	436	-0.0434	-7471B	PPB		02 Apr 2019 16:22:52	U	SMPL
CCV15 - 1	CCV15	32579	5.1740	-7471B	PPB		02 Apr 2019 16:25:08	U	SMPL
CCB15 - 1	CCB15	90	-0.0996	-7471B	PPB		02 Apr 2019 16:27:23	U	SMPL
K2179-11 - 1	TP-8S	441	-0.0426	-7471B	PPB		02 Apr 2019 16:29:39	U	SMPL
K2179-15 - 1	TP-8D	243	-0.0748	-7471B	PPB		02 Apr 2019 16:31:54	U	SMPL
K2179-19 - 1	TP-5S	1535	0.1349	-7471B	PPB		02 Apr 2019 16:34:10	U	SMPL
K2180-03 - 1	TP-7	34002	5.4050	-7471B	PPB		02 Apr 2019 16:36:26	U	SMPL
K2180-07 - 1	TP-6S	5384	0.7597	-7471B	PPB		02 Apr 2019 16:38:42	U	SMPL
K2180-11 - 1	TP-6D	2020	0.2137	-7471B	PPB		02 Apr 2019 16:40:59	U	SMPL
K2180-15 - 1	TP-9	6555	0.9498	-7471B	PPB		02 Apr 2019 16:43:15	U	SMPL
K2180-19 - 1	TP-10	12262	1.8762	-7471B	PPB		02 Apr 2019 16:45:32	U	SMPL
K2180-23 - 1	TP-10D	12595	1.9302	-7471B	PPB		02 Apr 2019 16:47:49	U	SMPL
K2180-27 - 1	TP-1	13679	2.1062	-7471B	PPB		02 Apr 2019 16:50:05	U	SMPL
CCV16 - 1	CCV16	32605	5.1783	-7471B	PPB		02 Apr 2019 16:52:21	U	SMPL
CCB16 - 1	CCB16	317	-0.0628	-7471B	PPB		02 Apr 2019 16:54:37	U	SMPL
K2180-31 - 1	TP-3	1909	0.1957	-7471B	PPB		02 Apr 2019 16:56:53	U	SMPL
K2182-01 - 1	S8A(2-10)COMP	1995	0.2096	-7471B	PPB		02 Apr 2019 16:59:09	U	SMPL
K2182-03 - 1	S8A-S8B(12-16)COMP	797	0.0152	-7471B	PPB		02 Apr 2019 17:01:25	U	SMPL
K2182-05 - 1	S8B(2-10)COMP	577	-0.0206	-7471B	PPB		02 Apr 2019 17:03:41	U	SMPL
K2182-07 - 1	S7C(2-10)COMP	776	0.0117	-7471B	PPB		02 Apr 2019 17:05:57	U	SMPL
K2182-09 - 1	S7C(12-20)COMP	1141	0.0710	-7471B	PPB		02 Apr 2019 17:08:13	U	SMPL
K2183-01 - 1	FK-0D015-03-002	811	0.0174	-7471B	PPB		02 Apr 2019 17:10:30	U	SMPL
K2184-01 - 1	FK-0D015-03-002	6008	0.8610	-7471B	PPB		02 Apr 2019 17:12:47	U	SMPL
PB118507BL - 1	PBS	335	-0.0598	-7471B	PPB		02 Apr 2019 17:15:04	U	SMPL
PB118507BS - 1	LCSS	21809	3.4258	-7471B	PPB		02 Apr 2019 17:17:20	U	SMPL
CCV17 - 1	CCV17	30867	4.8961	-7471B	PPB		02 Apr 2019 17:19:35	U	SMPL
CCB17 - 1	CCB17	313	-0.0634	-7471B	PPB		02 Apr 2019 17:21:51	U	SMPL
K1689-03 - 1	NB-08-032919-A	9658	1.4535	-7471B	PPB		02 Apr 2019 17:24:07	U	SMPL
K1689-03DUP - 1	NB-08-032919-ADUP	9731	1.4653	-7471B	PPB		02 Apr 2019 17:26:23	U	SMPL
K1689-03MS - 1	NB-08-032919-AMS	32115	5.0987	-7471B	PPB		02 Apr 2019 17:28:38	U	SMPL
K1689-03MSD - 1	NB-08-032919-AMSD	32188	5.1106	-7471B	PPB		02 Apr 2019 17:30:54	U	SMPL
K1689-05 - 1	NB-08-032919-B	8747	1.3056	-7471B	PPB		02 Apr 2019 17:33:11	U	SMPL
K1692-05 - 1	OK-03-032919	4412	0.6019	-7471B	PPB		02 Apr 2019 17:35:28	U	SMPL
K2056-01 - 1	CS-7	2564	0.3020	-7471B	PPB		02 Apr 2019 17:37:44	U	SMPL
K2056-02 - 1	CS-8	417	-0.0465	-7471B	PPB		02 Apr 2019 17:40:02	U	SMPL
K2056-03 - 1	CS-9	456	-0.0402	-7471B	PPB		02 Apr 2019 17:42:19	U	SMPL
K2056-04 - 1	CS-10	1856	0.1871	-7471B	PPB		02 Apr 2019 17:44:34	U	SMPL
CCV18 - 1	CCV18	30904	4.9021	-7471B	PPB		02 Apr 2019 17:46:50	U	SMPL
CCB18 - 1	CCB18	170	-0.0866	-7471B	PPB		02 Apr 2019 17:49:05	U	SMPL
K2171-01 - 1	GW-COMP	1759	0.1713	-7471B	PPB		02 Apr 2019 17:51:22	U	SMPL
K2173-01 - 1	T-3	161	-0.0881	-7471B	PPB		02 Apr 2019 17:53:37	U	SMPL
K2174-01 - 1	RBR200030	2089	0.2249	-7471B	PPB		02 Apr 2019 17:55:53	U	SMPL
K2178-01 - 1	COMP-16	9748	1.4681	-7471B	PPB		02 Apr 2019 17:58:09	U	SMPL
K2178-02 - 1	COMP-17	8388	1.2473	-7471B	PPB		02 Apr 2019 18:00:24	U	SMPL
K2178-03 - 1	COMP-18	18819	2.9405	-7471B	PPB		02 Apr 2019 18:02:41	U	SMPL
K2185-02 - 1	P-4-OW(0-10)	156641	25.3119	-7471B	PPB		02 Apr 2019 18:04:57	U	SMPL
K2186-01 - 1	TP-TRACK-89-90-SOIL	11693	1.7838	-7471B	PPB		02 Apr 2019 18:07:14	U	SMPL
K2186-05 - 1	TP-TRACK-86-87-SOIL	1261	0.0905	-7471B	PPB		02 Apr 2019 18:09:34	U	SMPL
K2186-09 - 1	BELT-PARKWAY-STAGING-YARD-1	7083	1.0355	-7471B	PPB		02 Apr 2019 18:11:50	U	SMPL
CCV19 - 1	CCV19	30492	4.8353	-7471B	PPB		02 Apr 2019 18:14:07	U	SMPL
CCB19 - 1	CCB19	156	-0.0889	-7471B	PPB		02 Apr 2019 18:16:23	U	SMPL
K2186-13 - 1	BELT-PARKWAY-STAGING-YARD-2	1222	0.0841	-7471B	PPB		02 Apr 2019 18:18:41	U	SMPL
K2186-17 - 1	BELT-PARKWAY-STAGING-YARD-3	5396	0.7617	-7471B	PPB		02 Apr 2019 18:20:57	U	SMPL
K2186-21 - 1	TP-TRACK-78-79	6091	0.8745	-7471B	PPB		02 Apr 2019 18:23:13	U	SMPL
K2179-03LX5 - 1		370	-0.0542	-7471B	PPB		02 Apr 2019 18:25:29	U	SMPL
K2179-03A - 1		23438	3.6903	-7471B	PPB		02 Apr 2019 18:27:45	U	SMPL
K1689-03LX5 - 1		1913	0.1963	-7471B	PPB		02 Apr 2019 18:30:01	U	SMPL
K1689-03A - 1		32330	5.1336	-7471B	PPB		02 Apr 2019 18:32:18	U	SMPL
K2185-02DLX2 - 1		32173	5.1081	-7471B	PPB		02 Apr 2019 18:34:35	U	SMPL
CCV20 - 1	CCV20	31476	4.9950	-7471B	PPB		02 Apr 2019 18:36:53	U	SMPL
CCB20 - 1	CCB20	149	-0.0900	-7471B	PPB		02 Apr 2019 18:39:09	U	SMPL