



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB101813

Ambient Barometric Pressure (mmHg): 760

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.0	80.6	80.6	04/02/2019	10:05
2	K1695-02		1	9	>100.0	>212.0	>212.0	04/02/2019	10:35
3	K1695-02DUP		1	9	>100.0	>212.0	>212.0	04/02/2019	10:50
4	K1696-01		1	7	>100.0	>212.0	>212.0	04/02/2019	11:20
5	K1696-02		1	6	>100.0	>212.0	>212.0	04/02/2019	11:45
6	K1696-03		1	7	>100.0	>212.0	>212.0	04/02/2019	12:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB101813

WorkList Name : FP4219

WorkList ID : 124272

Date : 04-02-2019 09:29:31

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/04/2019	Water	K1695-02	Flash Point	Cool 4 deg C	PSEG01	K31	SV-28399L	04/01/2019	1010A
04/08/2019	Water	K1696-01	Flash Point	Cool 4 deg C	PSEG01	K31	COMP-1	04/01/2019	1010A
04/08/2019	Water	K1696-02	Flash Point	Cool 4 deg C	PSEG01	K31	COMP-2	04/01/2019	1010A
04/08/2019	Water	K1696-03	Flash Point	Cool 4 deg C	PSEG01	K31	21860	04/01/2019	1010A

Date/Time 04/02/19 09:30
 Received by: KS
 Relinquished by: CP

Date/Time 04/02/19 12:15
 Received by: CP
 Relinquished by: KS