



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB101819

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	04/02/2019	17:30
2	CAL2	1	Water	NA	NA	20.2	7.00	04/02/2019	17:31
3	CAL3	1	Water	NA	NA	20.2	10.02	04/02/2019	17:32
4	ICV	1	Water	NA	NA	20.1	6.99	04/02/2019	17:33
5	CCV1	1	Water	NA	NA	21.1	2.02	04/02/2019	17:34
6	K2197-02	1	Solid	20.02	20	20.8	8.84	04/02/2019	17:35
7	K2197-02DUP	1	Solid	20.03	20	20.9	8.85	04/02/2019	17:36
8	K2197-04	1	Solid	20.04	20	20.8	8.70	04/02/2019	17:37
9	K2199-02	1	Solid	20.02	20	20.5	8.88	04/02/2019	17:38
10	K2199-04	1	Solid	20.03	20	20.4	8.34	04/02/2019	17:40
11	K2203-02	1	Solid	20.02	20	21.0	7.96	04/02/2019	17:41
12	K2203-04	1	Solid	20.04	20	20.8	7.31	04/02/2019	17:43
13	K2214-04	1	Solid	20.02	20	21.4	7.55	04/02/2019	17:45
14	K2214-08	1	Solid	20.03	20	21.7	6.63	04/02/2019	17:46
15	K2221-02	1	Solid	20.04	20	22.1	5.62	04/02/2019	17:48
16	CCV2	1	Water	NA	NA	20.1	12.02	04/02/2019	17:50
17	K2221-04	1	Solid	20.05	20	22.1	8.62	04/02/2019	18:00
18	CCV3	1	Water	NA	NA	20.2	2.02	04/02/2019	18:02

20101819

WORKLIST(Hardcopy Internal Chain)

Date : 04-02-2019 16:24:26

WorkList ID : 124300

WorkList Name : CORROSIONITY K2197

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/02/2019	Solid	K2197-02	Corrosivity	Cool 4 deg C	PSEG05	K31	CL-02-040219-A	04/02/2019	9045D
04/02/2019	Solid	K2197-04	Corrosivity	Cool 4 deg C	PSEG05	K31	CL-02-040219-B	04/02/2019	9045D
04/02/2019	Solid	K2199-02	Corrosivity	Cool 4 deg C	PSEG05	K31	OR-02-040219-A	04/02/2019	9045D
04/02/2019	Solid	K2199-04	Corrosivity	Cool 4 deg C	PSEG05	K31	OR-02-040219-B	04/02/2019	9045D
04/02/2019	Solid	K2203-02	Corrosivity	Cool 4 deg C	PSEG05	K31	SU-03-040219-A	04/02/2019	9045D
04/02/2019	Solid	K2203-04	Corrosivity	Cool 4 deg C	PSEG05	K31	SU-03-040219-B	04/02/2019	9045D
04/02/2019	Solid	K2214-04	Corrosivity	Cool 4 deg C	PSEG01	K31	TP-2-S	04/02/2019	9045D
04/02/2019	Solid	K2214-08	Corrosivity	Cool 4 deg C	PSEG01	K31	TP-2-D	04/02/2019	9045D
04/02/2019	Solid	K2221-02	Corrosivity	Cool 4 deg C	SKAN04	F51	S10B(2-10)COMP	04/02/2019	9045D
04/02/2019	Solid	K2221-04	Corrosivity	Cool 4 deg C	SKAN04	F51	S2B(2-10)COMP	04/02/2019	9045D

Date/Time 04-02-19 16:25
 Received by: SP
 Relinquished by: SP

Date/Time 04-02-19 16:30
 Received by: SP
 Relinquished by: SP