

TOTAL VOLATILE SOLIDS 160.4

TEMP1 IN: 103 04/02/2019 15:00 TEMP OUT: 105 04/03/2019 07:00
 TEMP2 IN: 105 04/03/2019 08:00 TEMP OUT: 105 04/03/2019 09:35
 TEMP3 IN: 550 04/03/2019 10:40 TEMP OUT: 550 04/03/2019 12:15
 TEMP3 IN: 550 04/03/2019 13:20 TEMP OUT: 550 04/03/2019 15:00

Run Number: LB101831

SUPERVISOR: apatel

ANALYST: Aparna

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Empty Dish + Sample Weight (g)	1st Dish + SampleWt Drying @103-@105°C (g)	Final Dish + SampleWt Drying @103-@105°C (g)	Dish + Samplewt Drying @550 (±50) °C (g)	Final Dish + Samplewt Drying @550 (±50) °C (g)	Weight Diff (g)	Result %
1	LB101831BL	59.6231	59.6231	59.6231	59.6231	59.6231	59.6231	59.6231	0	0
2	K2213-01	60.801	60.801	84.747	83.0827	83.0827	82.7613	82.7613	0.3214	1.4
3	K2213-01DUP	59.3983	59.3983	83.5186	81.841	81.841	81.6924	81.6924	0.1486	0.7
4	K2213-03	59.9823	59.9823	82.9992	81.598	81.598	81.4624	81.4624	0.1356	0.6
5	K2213-05	59.5637	59.5637	83.6417	82.368	82.368	82.2099	82.2099	0.1581	0.7
6	K2213-07	95.096	95.096	119.118	117.8045	117.8045	117.6145	117.6145	0.19	0.8
7	K2221-01	68.8398	68.8398	91.841	89.745	89.745	89.5038	89.5038	0.2412	1.2
8	K2221-03	83.4295	83.4295	106.4159	105.793	105.793	105.597	105.597	0.196	0.9

A = Sample Weight (g)

B = Final Dish + Samplewt Drying @550 (±50) °C (g)

C = Final Dish + SampleWt Drying @103-@105°C (g)

D = Weight (g)

E = Final Empty Dish Weight (g)

F = Final Dish + SampleWt Drying @103-@105°C (g)

Weight D = C - B

Result % = $\frac{D}{F - E} \star 100$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TVS040219

WorkList ID : 124247

Date : 04-02-2019 07:57:02

20101831

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/08/2019	Solid	K2213-01	TVS	Cool 4 deg C	SKAN04	F52	S9A(2-10)COMP	04/01/2019	160.4
04/08/2019	Solid	K2213-03	TVS	Cool 4 deg C	SKAN04	F52	S9A(12)-S9B(4)-S9B(8)-S9B(1	04/01/2019	160.4
04/08/2019	Solid	K2213-05	TVS	Cool 4 deg C	SKAN04	F52	S9C(2-10)COMP	04/01/2019	160.4
04/08/2019	Solid	K2213-07	TVS	Cool 4 deg C	SKAN04	F52	S10A(2-10)COMP	04/01/2019	160.4
04/09/2019	Solid	K2221-01	TVS	Cool 4 deg C	SKAN04	F51	S10B(2-10)COMP	04/02/2019	160.4
04/09/2019	Solid	K2221-03	TVS	Cool 4 deg C	SKAN04	F51	S2B(2-10)COMP	04/02/2019	160.4

Date/Time 04/02/19 14:30

Received by: AS

Relinquished by: CS

Date/Time 04/02/19 15:40

Received by: AS

Relinquished by: AS