


LB101832

ISM02.4

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.1577e-004

C= -3.0637e-002

Rho= 0.9992458

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	% D
0.00	0.000	0.002	0.002	153	0.000	153					—
0.2	0.200	0.181	-0.019	982	0.0 %	982					-10
2.5	2.500	2.358	-0.142	11072	0.0 %	11072					-6
5.0	5.000	5.182	0.182	24157	0.0 %	24157					4
7.5	7.500	7.674	0.174	35709	0.0 %	35709					2
10.0	10.000	9.802	-0.198	45568	0.0 %	45568					-2

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	153	-		0 ISM02.4	PPB	4/3/2019 12:05	Std
	0.2 S0.2	982	-		0.2 ISM02.4	PPB	4/3/2019 12:07	Std
	2.5 S2.5	11072	-		2.5 ISM02.4	PPB	4/3/2019 12:09	Std
	5 S5.0	24157	-		5 ISM02.4	PPB	4/3/2019 12:11	Std
	7.5 S7.5	35709	-		7.5 ISM02.4	PPB	4/3/2019 12:14	Std
	10 S10.0	45568	-		10 ISM02.4	PPB	4/3/2019 12:22	Std
ICV	ICV	17025	3.6429	-	ISM02.4	PPB	4/3/2019 12:26	SMPL
ICB	ICB	79	-0.0136	-	ISM02.4	PPB	4/3/2019 12:28	SMPL
CCV025	CCV025	22375	4.7973	-	ISM02.4	PPB	4/3/2019 12:31	SMPL
CCB025	CCB025	25	-0.0252	-	ISM02.4	PPB	4/3/2019 12:33	SMPL
PB118570BL	PBW570	65	-0.0166	-	ISM02.4	PPB	4/3/2019 12:35	SMPL
K2028-01	MA41D5	86	-0.0121	-	ISM02.4	PPB	4/3/2019 12:37	SMPL
K2028-02	MA41E6	36	-0.0229	-	ISM02.4	PPB	4/3/2019 12:40	SMPL
K2028-03	MA41F7	76	-0.0142	-	ISM02.4	PPB	4/3/2019 12:42	SMPL
K2028-04	MA41G5	45	-0.0209	-	ISM02.4	PPB	4/3/2019 12:44	SMPL
K2028-05	MA41H4	49	-0.0201	-	ISM02.4	PPB	4/3/2019 12:47	SMPL
K2028-06	MA41H6	29	-0.0244	-	ISM02.4	PPB	4/3/2019 12:49	SMPL
K2028-07	MA41J1	36	-0.0229	-	ISM02.4	PPB	4/3/2019 12:51	SMPL
K2028-11	MA41E2	34	-0.0233	-	ISM02.4	PPB	4/3/2019 12:53	SMPL
K2028-12	MA41H7	11	-0.0283	-	ISM02.4	PPB	4/3/2019 12:56	SMPL
K2028-13	MA41H8	40	-0.022	-	ISM02.4	PPB	4/3/2019 12:58	SMPL
K2028-14	MA41H8D	51	-0.0196	-	ISM02.4	PPB	4/3/2019 13:00	SMPL
K2028-15	MA41H8S	5123	1.0748	-	ISM02.4	PPB	4/3/2019 13:02	SMPL
K2028-16	MA41H9	42	-0.0216	-	ISM02.4	PPB	4/3/2019 13:05	SMPL
K2028-17	MA41K7	122	-0.0043	-	ISM02.4	PPB	4/3/2019 13:07	SMPL
K2028-18	MA4196	83	-0.0127	-	ISM02.4	PPB	4/3/2019 13:09	SMPL
PB118571BL	PBW571	81	-0.0132	-	ISM02.4	PPB	4/3/2019 13:11	SMPL
K2123-01	MC0B36	75	-0.0145	-	ISM02.4	PPB	4/3/2019 13:14	SMPL
K2123-02	MC0B41	135	-0.0015	-	ISM02.4	PPB	4/3/2019 13:16	SMPL
K2123-03	MC0B42	65	-0.0166	-	ISM02.4	PPB	4/3/2019 13:18	SMPL
CCV026	CCV026	22586	4.8428	-	ISM02.4	PPB	4/3/2019 13:21	SMPL
CCB026	CCB026	28	-0.0246	-	ISM02.4	PPB	4/3/2019 13:23	SMPL
K2123-04	MC0B43	118	-0.0052	-	ISM02.4	PPB	4/3/2019 13:25	SMPL
K2123-05	MC0B44	65	-0.0166	-	ISM02.4	PPB	4/3/2019 13:27	SMPL
K2123-06	MC0B45	85	-0.0123	-	ISM02.4	PPB	4/3/2019 13:30	SMPL
K2123-07	MC0B46	106	-0.0078	-	ISM02.4	PPB	4/3/2019 13:32	SMPL
K2123-08	MC0B37	95	-0.0101	-	ISM02.4	PPB	4/3/2019 13:34	SMPL
K2123-09	MC0B38	76	-0.0142	-	ISM02.4	PPB	4/3/2019 13:36	SMPL
K2123-10	MC0B39	66	-0.0164	-	ISM02.4	PPB	4/3/2019 13:39	SMPL
K2123-11	MC0B40	48	-0.0203	-	ISM02.4	PPB	4/3/2019 13:41	SMPL
K2123-12	MC0B47	45	-0.0209	-	ISM02.4	PPB	4/3/2019 13:43	SMPL
K2123-13	MC0B49	33	-0.0235	-	ISM02.4	PPB	4/3/2019 13:45	SMPL
K2123-14	MC0B51	42	-0.0216	-	ISM02.4	PPB	4/3/2019 13:48	SMPL
K2123-15	MC0B52	55	-0.0188	-	ISM02.4	PPB	4/3/2019 13:50	SMPL
K2123-16	MC0B53	24	-0.0255	-	ISM02.4	PPB	4/3/2019 13:52	SMPL
K2123-17	MC0B48	263	0.0261	-	ISM02.4	PPB	4/3/2019 13:55	SMPL

LB101832
INSTRUMENT ID : CV1

K2123-18	MCOB50	47	-0.0205 -	ISM02.4	PPB	4/3/2019 13:57	SMPL
K2123-19	MCOB54	37	-0.0227 -	ISM02.4	PPB	4/3/2019 13:59	SMPL
K2123-20	MCOB54D	23	-0.0257 -	ISM02.4	PPB	4/3/2019 14:01	SMPL
K2123-21	MCOB54S	4956	1.0387 -	ISM02.4	PPB	4/3/2019 14:04	SMPL
CCV027	CCV027	22216	4.763 -	ISM02.4	PPB	4/3/2019 14:06	SMPL
CCB027	CCB027	-32	-0.0375 -	ISM02.4	PPB	4/3/2019 14:08	SMPL
