



TOTAL Dissolved Solids - SM2540C

Run Number: LB101833

SUPERVISOR: apatel

ANALYST: Aparna

Date: 04/03/2019

BalanceID: WC SC-3

Filter ID: 14466061

TEMP1 IN:	<u>103</u>	<u>04/02/2019</u>	<u>15:05</u>	TEMP OUT:	<u>103</u>	<u>04/03/2019</u>	<u>07:00</u>	OvenID1:	<u>WC Oven-2</u>	Thermo ID1:	<u>WET-OVEN#2</u>
TEMP2 IN:	<u>182</u>	<u>04/03/2019</u>	<u>07:05</u>	TEMP OUT:	<u>182</u>	<u>04/03/2019</u>	<u>08:45</u>	OvenID2:	<u>WC-OVEN-3</u>	Thermo ID2:	<u>WET-OVEN#3</u>
TEMP3 IN:	<u>182</u>	<u>04/03/2019</u>	<u>10:00</u>	TEMP OUT:	<u>182</u>	<u>04/03/2019</u>	<u>11:45</u>	OvenID3:	<u>WC-OVEN-3</u>	Thermo ID3:	<u>WET-OVEN#3</u>

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @180 (±2) °C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @180 (±2) °C (g)	Final Empty Dish+Sample weight after 1.5hr drying @180 (±2) °C (g)	Weight (g)	Result mg/L
1	LB101833BL	LB101833BL	94.2265	94.2265	100	94.2265	94.2265	94.2265	0.0000	0
2	LB101833BS	LB101833BS	90.1235	90.1235	100	90.1340	90.1340	90.1340	0.0105	105
3	K2233-02	FE-19C	94.2768	94.2768	100	94.6460	94.6460	94.6460	0.3692	3692
4	K2233-02DUP	FE-19CDUP	78.0845	78.0845	100	78.4619	78.4619	78.4619	0.3774	3774

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = 2nd Empty Dish + Sample weight after 1.5 hr drying @180 (±2) °C (g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)


WorkList Name : TDS040219

WorkList ID : 124296

Date : 04-02-2019 15:21:53

KB101833

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/09/2019	Water	K2233-02	TDS	Cool 4 deg C	M&MM01	F51	FE-19C	04/02/2019	SM2540 C

Date/Time 04/02/19 14:30
 Received by: 
 Relinquished by: 

Date/Time 04/02/19 15:40
 Received by: 
 Relinquished by: 