

LB101837


7471B

INSTRUMENT ID: CU1

Linear

 μ Abs.:

57314



A= 0.0000e+000

B= 1.7337e-004

C= -3.8588e-002

Rho= 0.9997785

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.024	-0.024	83	0.000	83				
0.2	0.200	0.166	-0.034	1178	0.0 %	1178				
2.5	2.500	2.562	0.062	14999	0.0 %	14999				
5.0	5.000	4.962	-0.038	28842	0.0 %	28842				
7.5	7.500	7.637	0.137	44272	0.0 %	44272				
10.0	10.000	9.898	-0.102	57314	0.0 %	57314				

LB101837 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 03 Apr 2019 14:31:38

Sample ID	Extended ID	μ Abs.	Conc	Sind Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	83	-	0.000	7471B	PPB	03 Apr 2019 14:38:09	S	Std
0.2 - 1	50.2	1178	-	0.200	7471B	PPB	03 Apr 2019 14:40:27	S	Std
2.5 - 1	52.5	14999	-	2.500	7471B	PPB	03 Apr 2019 14:42:44	S	Std
5.0 - 1	55.0	28842	-	5.000	7471B	PPB	03 Apr 2019 14:45:01	S	Std
7.5 - 1	57.5	44272	-	7.500	7471B	PPB	03 Apr 2019 14:47:17	S	Std
10.0 - 1	570.0	57314	-	10.000	7471B	PPB	03 Apr 2019 14:57:37	S	Std
ICV86 - 1	ICV86	22639	3.8864	-	7471B	PPB	03 Apr 2019 15:02:03	U	SMPL
ICB86 - 1	ICB86	9	-0.0370	-	7471B	PPB	03 Apr 2019 15:04:19	U	SMPL
CCV28 - 1	CCV28	28937	4.9783	-	7471B	PPB	03 Apr 2019 15:06:35	U	SMPL
CCB28 - 1	CCB28	13	-0.0363	-	7471B	PPB	03 Apr 2019 15:08:50	U	SMPL
CRA - 1	CRA	1153	0.1613	-	7471B	PPB	03 Apr 2019 15:11:05	U	SMPL
HighStd - 1	HighStd	56298	9.7219	-	7471B	PPB	03 Apr 2019 15:13:20	U	SMPL
ChkStd - 1	ChkStd	40257	6.9408	-	7471B	PPB	03 Apr 2019 15:15:36	U	SMPL
PB118563BL - 1	PBS	-59	-0.0488	-	7471B	PPB	03 Apr 2019 15:17:53	U	SMPL
PB118563BS - 1	LCSS	20799	3.5674	-	7471B	PPB	03 Apr 2019 15:20:11	U	SMPL
K1695-04 - 1	DPC-91723-SO-TP5-A	1102	0.1525	-	7471B	PPB	03 Apr 2019 15:22:28	U	SMPL
K1695-04DUP - 1	DPC-91723-SO-TP5-ADUP	1081	0.1488	-	7471B	PPB	03 Apr 2019 15:24:44	U	SMPL
K1695-04MS - 1	DPC-91723-SO-TP5-AMS	23668	4.0648	-	7471B	PPB	03 Apr 2019 15:27:01	U	SMPL
K1695-04MSD - 1	DPC-91723-SO-TP5-AMSD	23945	4.1128	-	7471B	PPB	03 Apr 2019 15:29:19	U	SMPL
K1695-08 - 1	DPC-91723-SO-TP5-B	1130	0.1573	-	7471B	PPB	03 Apr 2019 15:31:34	U	SMPL
CCV29 - 1	CCV29	29188	5.0218	-	7471B	PPB	03 Apr 2019 15:33:51	U	SMPL
CCB29 - 1	CCB29	-13	-0.0408	-	7471B	PPB	03 Apr 2019 15:36:07	U	SMPL
K2190-01 - 1	VN-SP-1	80159	13.8587	-	7471B	PPB	03 Apr 2019 15:38:24	U	SMPL
K2190-04 - 1	VN-SP-2	67267	11.6236	-	7471B	PPB	03 Apr 2019 15:40:39	U	SMPL
K2196-01 - 1	HD-01-040119	1029	0.1398	-	7471B	PPB	03 Apr 2019 15:42:57	U	SMPL
K2197-01 - 1	CL-02-040219-A	2059	0.3184	-	7471B	PPB	03 Apr 2019 15:45:14	U	SMPL
K2197-03 - 1	CL-02-040219-B	3173	0.5115	-	7471B	PPB	03 Apr 2019 15:47:30	U	SMPL
K2199-01 - 1	OR-02-040219-A	1418	0.2073	-	7471B	PPB	03 Apr 2019 15:49:46	U	SMPL
K2199-03 - 1	OR-02-040219-B	1401	0.2043	-	7471B	PPB	03 Apr 2019 15:52:03	U	SMPL
K2203-01 - 1	SU-03-040219-A	3643	0.5930	-	7471B	PPB	03 Apr 2019 15:54:20	U	SMPL
K2203-03 - 1	SU-03-040219-B	1577	0.2348	-	7471B	PPB	03 Apr 2019 15:56:37	U	SMPL
K2213-01 - 1	S9A(2-10)COMP	597	0.0649	-	7471B	PPB	03 Apr 2019 15:58:53	U	SMPL
CCV30 - 1	CCV30	29073	5.0018	-	7471B	PPB	03 Apr 2019 16:01:08	U	SMPL
CCB30 - 1	CCB30	11	-0.0367	-	7471B	PPB	03 Apr 2019 16:03:24	U	SMPL
K2213-03 - 1	S9A(12)-S9B(4)-S9B(8)-S9B(12)-S9C(12)	1563	0.2324	-	7471B	PPB	03 Apr 2019 16:05:40	U	SMPL
K2213-05 - 1	S9C(2-10)COMP	1701	0.2563	-	7471B	PPB	03 Apr 2019 16:07:55	U	SMPL
K2213-07 - 1	S10A(2-10)COMP	844	0.1077	-	7471B	PPB	03 Apr 2019 16:10:11	U	SMPL
K2214-03 - 1	TP-2-S	7289	1.2251	-	7471B	PPB	03 Apr 2019 16:12:26	U	SMPL
K2214-07 - 1	TP-2-D	666	0.0769	-	7471B	PPB	03 Apr 2019 16:14:42	U	SMPL
K2219-01 - 1	TP-1-2-COMP	13009	2.2168	-	7471B	PPB	03 Apr 2019 16:16:59	U	SMPL
K2219-02 - 1	TP-2-3-COMP	10851	1.8427	-	7471B	PPB	03 Apr 2019 16:19:15	U	SMPL
PB118564BL - 1	PBS	25	-0.0343	-	7471B	PPB	03 Apr 2019 16:21:33	U	SMPL
PB118564BS - 1	LCSS	20887	3.5826	-	7471B	PPB	03 Apr 2019 16:23:50	U	SMPL
K2221-01 - 1	S10B(2-10)COMP	1661	0.2494	-	7471B	PPB	03 Apr 2019 16:26:06	U	SMPL
CCV31 - 1	CCV31	30194	5.1962	-	7471B	PPB	03 Apr 2019 16:28:22	U	SMPL
CCB31 - 1	CCB31	14	-0.0362	-	7471B	PPB	03 Apr 2019 16:30:38	U	SMPL
K2221-01DUP - 1	S10B(2-10)COMP	1701	0.2563	-	7471B	PPB	03 Apr 2019 16:32:54	U	SMPL
K2221-01MS - 1	S10B(2-10)COMPMS	23664	4.0641	-	7471B	PPB	03 Apr 2019 16:35:09	U	SMPL
K2221-01MSD - 1	S10B(2-10)COMPMSD	23459	4.0285	-	7471B	PPB	03 Apr 2019 16:37:26	U	SMPL
K2221-03 - 1	S2B(2-10)COMP	3527	0.5729	-	7471B	PPB	03 Apr 2019 16:39:41	U	SMPL
K2222-01 - 1	DRUM-1-4	72813	12.5851	-	7471B	PPB	03 Apr 2019 16:41:57	U	SMPL
K2227-01 - 1	CONCRETE-PILE-1-A	180	-0.0074	-	7471B	PPB	03 Apr 2019 16:44:13	U	SMPL
K2227-03 - 1	CONCRETE-PILE-1-B	189	-0.0058	-	7471B	PPB	03 Apr 2019 16:46:32	U	SMPL
K2227-05 - 1	CONCRETE-PILE-1-C	124	-0.0171	-	7471B	PPB	03 Apr 2019 16:48:49	U	SMPL
K2227-07 - 1	CONCRETE-PILE-1-D	388	0.0287	-	7471B	PPB	03 Apr 2019 16:51:06	U	SMPL
K1695-04LX5 - 1		170	-0.0091	-	7471B	PPB	03 Apr 2019 16:53:22	U	SMPL
CCV32 - 1	CCV32	31082	5.3501	-	7471B	PPB	03 Apr 2019 16:55:38	U	SMPL
CCB32 - 1	CCB32	9	-0.0370	-	7471B	PPB	03 Apr 2019 16:57:53	U	SMPL
K1695-04A - 1		24073	4.1350	-	7471B	PPB	03 Apr 2019 17:00:09	U	SMPL
K2221-01LX5 - 1		236	0.0023	-	7471B	PPB	03 Apr 2019 17:02:25	U	SMPL
K2221-01A - 1		23769	4.0823	-	7471B	PPB	03 Apr 2019 17:04:41	U	SMPL
K2190-01DLX2 - 1		40565	6.9942	-	7471B	PPB	03 Apr 2019 17:06:57	U	SMPL
K2190-04DLX2 - 1		32876	5.6612	-	7471B	PPB	03 Apr 2019 17:09:13	U	SMPL
K2222-01DLX2 - 1		36714	6.3266	-	7471B	PPB	03 Apr 2019 17:11:30	U	SMPL
CCV33 - 1	CCV33	30747	5.2921	-	7471B	PPB	03 Apr 2019 17:13:48	U	SMPL
CCB33 - 1	CCB33	9	-0.0370	-	7471B	PPB	03 Apr 2019 17:16:05	U	SMPL