



TOTAL SOLIDS - SM2540B

SUPERVISOR: apatelANALYST: AparnaDate: 04/04/2019OVEN TEMP1: 103 °CTEMP: 105 °CDATE & TIME IN: 04/03/2019 15:45 OUT: 04/04/2019 07:00OVEN TEMP2: 105 °CTEMP: 105 °CDATE & TIME IN: 04/04/2019 08:05 OUT: 04/04/2019 09:35Run Number: LB101843BalanceID: WC SC-3OvenID: WC-OVEN-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB101843BL	LB101843BL	89.2365	89.2365	100	89.2365	89.2365	89.2365	0.0000	0
2	K2242-05	TWP-02	90.0084	90.0084	100	91.6245	91.6245	91.6245	1.6161	16161
3	K2242-05DUP	TWP-02DUP	90.0640	90.0640	100	91.6554	91.6554	91.6554	1.5914	15914

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TS040319

WorkList ID : 124358

Date : 04-03-2019 15:11:24

12101843

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/09/2019	Water	K2242-05	TS	Cool 4 deg C	LIR001	F31	TWP-02	04/02/2019	SM2540 B

Date/Time

04/03/19 15:15

Received by:

AS

Relinquished by:

CS

Date/Time

04/03/19 16:00

Received by:

AS

Relinquished by:

AS