



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 4/4/2019

OVENTEMP IN Celsius(°C): 108
Time IN: 17:30
In Date: 04/03/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:20
Out Date: 04/04/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-1
Thermometer ID: %SOLIDS-OVEN

QC:LB101851

Lab ID	Client Sample ID	Dish#	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish + Dry Sample Wt(g) (C)	% Solid
K2193-01	HR-06-040319-A	1	1.14	9.67	8.76	89.3
K2193-03	HR-06-040319-B	2	1.17	9.88	8.88	88.5
K2193-05	HR-06-040319-C	3	1.17	9.85	9.02	90.4
K2195-01	NB-07-040319	4	1.15	9.95	9.08	90.1
K2230-01	RICH-REC-ROAD-STONE-14	5	1.14	9.75	9.61	98.4
K2230-02	RICH-REC-ROAD-STONE-14	6	1.13	9.7	9.53	98.0
K2231-01	RICH-REC-ROAD-STONE-13	7	1.14	9.98	9.75	97.4
K2231-02	RICH-REC-ROAD-STONE-13	8	1.11	9.55	9.33	97.4
K2232-01	RICH-REC-ROAD-STONE-12	9	1.18	9.5	9.00	94.0
K2232-02	RICH-REC-ROAD-STONE-12	10	1.14	9.81	9.07	91.5
K2234-01	SSSI-0401-S-DH-3	11	1.16	9.96	7.99	77.6
K2234-02	SSSI-0401-S-DH-3	12	1.14	9.94	7.25	69.4
K2239-01	390327721	13	1.15	9.95	1.45	3.4
K2240-01	SB-01-(0-2)	14	1.13	9.51	8.52	88.2
K2240-02	SB-01-(13.5-14)	15	1.14	9.97	9.22	91.5
K2240-03	SB-01-COMP	16	1.17	9.75	8.8	88.9
K2240-04	SB-02-(0-2)	17	1.14	9.94	9.08	90.2
K2240-05	SB-02-(0-2)-DUP	18	1.19	9.71	8.59	86.9
K2240-06	SB-02-(13.5-14)	19	1.15	9.95	9.08	90.1
K2240-07	SB-02-COMP	20	1.13	9.58	7.93	80.5
K2240-08	SB-03-(0-6)	21	1.17	9.98	8.82	86.8
K2240-09	SB-03-COMP	22	1.14	9.91	9.16	91.4
K2244-01	FK-EF-01-007-ABCD	23	1.15	9.66	8.56	87.1
K2244-02	FK-SF-01-007-ABCD	24	1.1	9.57	8.59	88.4
K2245-01	TP-1-S1	25	1.14	9.53	8.32	85.6
K2245-02	TP-1-S2	26	1.12	9.81	8.75	87.8
K2245-03	TP-1S	27	1.13	9.98	9.06	89.6
K2245-05	TP-1-D1	28	1.12	9.96	8.97	88.8
K2245-06	TP-1-D2	29	1.12	9.77	8.62	86.7
K2245-07	TP-1D	30	1.14	9.75	8.36	83.9
K2245-09	TP-3-S1	31	1.1	9.68	8.72	88.8
K2245-10	TP-3-S2	32	1.14	9.98	9.26	91.9
K2245-11	TP-3S	33	1.14	9.73	8.83	89.5
K2245-13	TP-3-D1	34	1.14	9.7	8.79	89.4
K2245-14	TP-3-D2	35	1.13	9.72	9.00	91.6
K2245-15	TP-3D	36	1.14	9.95	9.55	95.5
K2257-01	SP-1-A	37	1.12	9.55	8.27	84.8
K2257-02	SP-1-A-2	38	1.12	9.55	8.27	84.8
K2257-03	SP-1-A-3	39	1.12	9.55	8.27	84.8

**PERCENT SOLID**

Supervisor: apatel
Analyst: JIGNESH
Date: 4/4/2019

OVENTEMP IN Celsius(°C): 108
Time IN: 17:30
In Date: 04/03/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:20
Out Date: 04/04/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-1
Thermometer ID: %SOLIDS-OVEN

QC:LB101851

Lab ID	Client Sample ID	Dish#	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish + Dry Sample Wt(g) (C)	% Solid
K2257-04	SP-1-A-4	40	1.12	9.55	8.27	84.8
K2257-05	SP-1-A-5	41	1.12	9.55	8.27	84.8
K2257-06	SP-1-A-6	42	1.12	9.55	8.27	84.8
K2257-07	SP-1-A-7	43	1.12	9.55	8.27	84.8
K2257-08	SP-1-A	44	1.12	9.55	8.27	84.8
K2257-09	SP-1-B	45	1.12	9.97	8.6	84.5
K2257-10	SP-1-B-2	46	1.12	9.97	8.6	84.5
K2257-11	SP-1-B-3	47	1.12	9.97	8.6	84.5
K2257-12	SP-1-B-4	48	1.12	9.97	8.6	84.5
K2257-13	SP-1-B-5	49	1.12	9.97	8.6	84.5
K2257-14	SP-1-B-6	50	1.12	9.97	8.6	84.5
K2257-15	SP-1-B-7	51	1.12	9.97	8.6	84.5
K2257-16	SP-1-B	52	1.12	9.97	8.6	84.5
K2257-17	SP-1-C	53	1.13	9.7	8.41	84.9
K2257-18	SP-1-C-2	54	1.13	9.7	8.41	84.9
K2257-19	SP-1-C-3	55	1.13	9.7	8.41	84.9
K2257-20	SP-1-C-4	56	1.13	9.7	8.41	84.9
K2257-21	SP-1-C-5	57	1.13	9.7	8.41	84.9
K2257-22	SP-1-C-6	58	1.13	9.7	8.41	84.9
K2257-23	SP-1-C-7	59	1.13	9.7	8.41	84.9
K2257-24	SP-1-C	60	1.13	9.7	8.41	84.9
K2257-25	SP-1-D	61	1.14	9.88	8.4	83.1
K2257-26	SP-1-D-2	62	1.14	9.88	8.4	83.1
K2257-27	SP-1-D-3	63	1.14	9.88	8.4	83.1
K2257-28	SP-1-D-4	64	1.14	9.88	8.4	83.1
K2257-29	SP-1-D-5	65	1.14	9.88	8.4	83.1
K2257-30	SP-1-D-6	66	1.14	9.88	8.4	83.1
K2257-31	SP-1-D-7	67	1.14	9.88	8.4	83.1
K2257-32	SP-1-D	68	1.14	9.88	8.4	83.1
K2257-33	RO-20	69	1.00	2.00	2.00	100.0
K2258-01	W-OIL-001-A-E	70	1.00	2.00	2.00	100.0
K2258-02	STONE	71	1.00	2.00	2.00	100.0
K2258-04	OIL-DEBRIS-001-A-C	72	1.00	2.00	2.00	100.0

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$