

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB101868

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP70044
calibration std. phosphate 1 ppm	WP73304
calibration std. phosphate 0.5 ppm	WP73303
calibration std. phosphate 0.3 ppm	WP73302
calibration std. 0 ppm	WP73299
phosphate ICV std.	WP73306
phosphate CCV std.	WP73305
calibration std. phosphate 0.1 ppm	WP73301
calibration std. phosphate 0.05 ppm	WP73300
Combined reagent	WP73308
11 N sulfuric acid	WP71598
Phenolphthalein indicator	WP71599
Sodium hydroxide, 1N	WP71600

Intercept: -0.001Slope: 0.6513Regression: 0.999831

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	04/04/2019	12:08
2	CAL2	0.05	1	50	50	0.033	04/04/2019	12:08
3	CAL3	0.10	1	50	50	0.065	04/04/2019	12:09
4	CAL4	0.30	1	50	50	0.186	04/04/2019	12:09
5	CAL5	0.50	1	50	50	0.330	04/04/2019	12:10
6	CAL6	1.00	1	50	50	0.650	04/04/2019	12:11



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB101868

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.326	0.502	04/04/2019	12:13
2	ICB		1	50	50	0.000	0.002	04/04/2019	12:14
3	CCV1	0.50	1	50	50	0.327	0.504	04/04/2019	12:15
4	CCB1		1	50	50	0.000	0.002	04/04/2019	12:15
5	PB118615BL		1	50	50	0.000	0.002	04/04/2019	12:16
6	PB118615BS	0.50	1	50	50	0.308	0.474	04/04/2019	12:16
7	K2233-01		1	50	50	0.870	1.337	04/04/2019	12:17
8	K2256-13		1	50	50	0.213	0.329	04/04/2019	12:18
9	K2256-13DUP		1	50	50	0.211	0.326	04/04/2019	12:18
10	K2256-13MS	0.50	1	50	50	0.496	0.763	04/04/2019	12:19
11	K2256-13MSD	0.50	1	50	50	0.490	0.754	04/04/2019	12:20
12	K2233-01		5	50	50	0.317	0.488	04/04/2019	12:31
13	CCV2	0.50	1	50	50	0.326	0.502	04/04/2019	12:32
14	CCB2		1	50	50	0.000	0.002	04/04/2019	12:32