



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB101874

Ambient Barometric Pressure (mmHg): 755

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.0	80.6	80.9	04/04/2019	09:45
2	K2242-05		1	7	>100.0	>212.0	>212.0	04/04/2019	10:20
3	K2242-05DUP		1	7	>100.0	>212.0	>212.0	04/04/2019	10:50

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB101874

WorkList Name : FP4419

WorkList ID : 124391

Date : 04-04-2019 09:10:34

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/09/2019	Water	K2242-05	Flash Point	Cool 4 deg C	LIR001	F31	TWP-02	04/02/2019	1010A

Date/Time 04/04/19 09:15
 Received by: KS
 Relinquished by: CP

Date/Time 04/04/19 16:55
 Received by: CP
 Relinquished by: KS