

LB101875


7470A

INSTRUMENT 1D : CV1

Linear

 μ Abs.:

43190



A= 0.0000e+000

B= 2.2969e-004

C= -1.3575e-002

Rho= 0.9997238

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.005	-0.005	39	0.000	39				
0.2	0.200	0.169	-0.031	797	0.0 %	797				
2.5	2.500	2.428	-0.072	10630	0.0 %	10630				
5.0	5.000	5.172	0.172	22575	0.0 %	22575				
7.5	7.500	7.529	0.029	32837	0.0 %	32837				
10.0	10.000	9.907	-0.093	43190	0.0 %	43190				

LB101875 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 04 Apr 2019 13:36:47

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	S0-0	39	-	0.0000	7470A	PPB	04 Apr 2019 13:42:56	S	Std
0.2 - 1	S0-2	797	-	0.2000	7470A	PPB	04 Apr 2019 13:45:05	S	Std
2.5 - 1	S2-5	10630	-	2.5000	7470A	PPB	04 Apr 2019 13:47:13	S	Std
5.0 - 1	S5-0	22575	-	5.0000	7470A	PPB	04 Apr 2019 13:49:21	S	Std
7.5 - 1	S7-5	32837	-	7.5000	7470A	PPB	04 Apr 2019 13:51:28	S	Std
10.0 - 1	S10-0	43190	-	10.0000	7470A	PPB	04 Apr 2019 13:57:06	S	Std
ICV88 - 1	ICV88	17886	4.0946	-7470A	PPB	PPB	04 Apr 2019 14:00:31	U	SMPL
ICB88 - 1	ICB88	-25	-0.0193	-7470A	PPB	PPB	04 Apr 2019 14:02:37	U	SMPL
CCV40 - 1	CCV40	21475	4.9190	-7470A	PPB	PPB	04 Apr 2019 14:04:43	U	SMPL
CCB40 - 1	CCB40	-55	-0.0262	-7470A	PPB	PPB	04 Apr 2019 14:06:49	U	SMPL
CRA - 1	CRA	765	0.1621	-7470A	PPB	PPB	04 Apr 2019 14:08:55	U	SMPL
HighStd - 1	HighStd	42822	9.8222	-7470A	PPB	PPB	04 Apr 2019 14:11:01	U	SMPL
ChkStd - 1	ChkStd	30766	7.0530	-7470A	PPB	PPB	04 Apr 2019 14:13:07	U	SMPL
PB118601BL - 1	PBW	-63	-0.0280	-7470A	PPB	PPB	04 Apr 2019 14:15:14	U	SMPL
PB118601BS - 1	LCSW	15881	3.6341	-7470A	PPB	PPB	04 Apr 2019 14:17:23	U	SMPL
K2197-02 - 1	CL-02-040219-A	-17	-0.0175	-7470A	PPB	PPB	04 Apr 2019 14:19:30	U	SMPL
K2197-02DUP - 1	CL-02-040219-ADUP	15	-0.0101	-7470A	PPB	PPB	04 Apr 2019 14:21:38	U	SMPL
K2197-02MS - 1	CL-02-040219-AMS	17529	4.0127	-7470A	PPB	PPB	04 Apr 2019 14:23:45	U	SMPL
K2197-02MSD - 1	CL-02-040219-AMSD	17537	4.0145	-7470A	PPB	PPB	04 Apr 2019 14:25:53	U	SMPL
K2197-04 - 1	CL-02-040219-B	-26	-0.0195	-7470A	PPB	PPB	04 Apr 2019 14:27:59	U	SMPL
CCV41 - 1	CCV41	21652	4.9597	-7470A	PPB	PPB	04 Apr 2019 14:30:06	U	SMPL
CCB41 - 1	CCB41	-53	-0.0257	-7470A	PPB	PPB	04 Apr 2019 14:32:12	U	SMPL
K2199-02 - 1	OR-02-040219-A	4	-0.0127	-7470A	PPB	PPB	04 Apr 2019 14:34:19	U	SMPL
K2199-04 - 1	OR-02-040219-B	8	-0.0117	-7470A	PPB	PPB	04 Apr 2019 14:36:24	U	SMPL
K2203-02 - 1	SU-03-040219-A	-4	-0.0145	-7470A	PPB	PPB	04 Apr 2019 14:38:30	U	SMPL
K2203-04 - 1	SU-03-040219-B	13	-0.0106	-7470A	PPB	PPB	04 Apr 2019 14:40:37	U	SMPL
K2214-04 - 1	TP-2-S	18	-0.0094	-7470A	PPB	PPB	04 Apr 2019 14:42:43	U	SMPL
K2214-08 - 1	TP-2-D	32	-0.0062	-7470A	PPB	PPB	04 Apr 2019 14:44:51	U	SMPL
K2221-02 - 1	S10B(2-10)COMP	8	-0.0117	-7470A	PPB	PPB	04 Apr 2019 14:46:58	U	SMPL
K2221-04 - 1	S2B(2-10)COMP	5	-0.0124	-7470A	PPB	PPB	04 Apr 2019 14:49:05	U	SMPL
K2222-02 - 1	DRUM-1-4	24	-0.0081	-7470A	PPB	PPB	04 Apr 2019 14:51:13	U	SMPL
K2227-02 - 1	CONCRETE-PILE-1-A	19	-0.0092	-7470A	PPB	PPB	04 Apr 2019 14:53:19	U	SMPL
CCV42 - 1	CCV42	21500	4.9247	-7470A	PPB	PPB	04 Apr 2019 14:55:26	U	SMPL
CCB42 - 1	CCB42	-37	-0.0221	-7470A	PPB	PPB	04 Apr 2019 14:57:33	U	SMPL
K2227-04 - 1	CONCRETE-PILE-1-B	0	-0.0136	-7470A	PPB	PPB	04 Apr 2019 14:59:40	U	SMPL
K2227-06 - 1	CONCRETE-PILE-1-C	-5	-0.0147	-7470A	PPB	PPB	04 Apr 2019 15:01:46	U	SMPL
K2227-08 - 1	CONCRETE-PILE-1-D	5	-0.0124	-7470A	PPB	PPB	04 Apr 2019 15:03:53	U	SMPL
PB118624BL - 1	PBW	24	-0.0081	-7470A	PPB	PPB	04 Apr 2019 15:06:00	U	SMPL
PB118624BS - 1	LCSW	16351	3.7421	-7470A	PPB	PPB	04 Apr 2019 15:08:06	U	SMPL
K2222-06 - 1	VB-16206	970	0.2092	-7470A	PPB	PPB	04 Apr 2019 15:10:14	U	SMPL
K2222-06DUP - 1	VB-16206DUP	1023	0.2214	-7470A	PPB	PPB	04 Apr 2019 15:12:21	U	SMPL
K2222-06MS - 1	VB-16206MS	19579	4.4835	-7470A	PPB	PPB	04 Apr 2019 15:14:29	U	SMPL
K2222-06MSD - 1	VB-16206MSD	19166	4.3887	-7470A	PPB	PPB	04 Apr 2019 15:16:37	U	SMPL
K2242-01 - 1	TWP-02	745	0.1575	-7470A	PPB	PPB	04 Apr 2019 15:18:43	U	SMPL
CCV43 - 1	CCV43	21557	4.9378	-7470A	PPB	PPB	04 Apr 2019 15:20:50	U	SMPL
CCB43 - 1	CCB43	-28	-0.0200	-7470A	PPB	PPB	04 Apr 2019 15:22:57	U	SMPL
K2242-02 - 1	TWP-02-DUP	552	0.1132	-7470A	PPB	PPB	04 Apr 2019 15:25:04	U	SMPL
K2242-03 - 1	FIELD-BLANK	285	0.0519	-7470A	PPB	PPB	04 Apr 2019 15:27:10	U	SMPL
K2242-06 - 1	TWP-02	397	0.0776	-7470A	PPB	PPB	04 Apr 2019 15:29:16	U	SMPL
K2242-10 - 1	TWP-02-DUP	384	0.0746	-7470A	PPB	PPB	04 Apr 2019 15:31:24	U	SMPL
K2242-11 - 1	FIELD-BLANK	311	0.0579	-7470A	PPB	PPB	04 Apr 2019 15:33:31	U	SMPL
PB118601TB - 1	PB118547TB	18	-0.0094	-7470A	PPB	PPB	04 Apr 2019 15:35:38	U	SMPL
K2197-02LX5 - 1		34	-0.0058	-7470A	PPB	PPB	04 Apr 2019 15:37:46	U	SMPL
K2197-02A - 1		17745	4.0623	-7470A	PPB	PPB	04 Apr 2019 15:39:54	U	SMPL
K2222-06LX5 - 1		195	0.0312	-7470A	PPB	PPB	04 Apr 2019 15:42:02	U	SMPL
K2222-06A - 1		19405	4.4435	-7470A	PPB	PPB	04 Apr 2019 15:44:10	U	SMPL
CCV44 - 1	CCV44	21641	4.9571	-7470A	PPB	PPB	04 Apr 2019 15:46:16	U	SMPL
CCB44 - 1	CCB44	-46	-0.0241	-7470A	PPB	PPB	04 Apr 2019 15:48:23	U	SMPL