



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB101877

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	04/04/2019	14:10
2	CAL2	1	Water	NA	NA	20.1	7.01	04/04/2019	14:11
3	CAL3	1	Water	NA	NA	20.1	10.03	04/04/2019	14:12
4	ICV	1	Water	NA	NA	20.1	7.00	04/04/2019	14:14
5	CCV1	1	Water	NA	NA	20.1	2.00	04/04/2019	14:15
6	K2265-01	1	Water	NA	NA	19.3	7.67	04/04/2019	14:16
7	K2265-01DUP	1	Water	NA	NA	19.4	7.68	04/04/2019	14:17
8	CCV2	1	Water	NA	NA	20.2	12.02	04/04/2019	14:20

LB 101877

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K2265

WorkList ID : 124419

Date : 04-04-2019 13:57:43

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/04/2019	Water	K2265-01	pH	Cool 4 deg C	PSEG01	K31	BW-04419	04/04/2019	9040C

Date/Time 04-04-19 14:00
Received by: SD
Relinquished by: MA

Date/Time 04-04-19 14:35
Received by: CS
Relinquished by: SD