



PERCENT SOLID

Supervisor: apatel  
Analyst: JIGNESH  
Date: 4/5/2019

OVENTEMP IN Celsius(°C): 108  
Time IN: 16:10  
In Date: 04/04/2019  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103  
Time OUT: 07:49  
Out Date: 04/05/2019  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-1  
Thermometer ID: %SOLIDS-OVEN

QC:LB101879

| Lab ID   | Client Sample ID | Dish# | Dish Wt(g)<br>(A) | Dish +<br>Sample Wt(g)<br>(B) | Dish + Dry<br>Sample Wt(g)<br>(C) | % Solid |
|----------|------------------|-------|-------------------|-------------------------------|-----------------------------------|---------|
| K2255-01 | BFC34            | 1     | 1.18              | 9.82                          | 8.09                              | 80.0    |
| K2255-02 | BFC35            | 2     | 1.18              | 9.8                           | 8.45                              | 84.3    |
| K2255-03 | BFC36            | 3     | 1.13              | 9.55                          | 7.12                              | 71.1    |
| K2255-04 | BFC46            | 4     | 1.13              | 9.59                          | 9.00                              | 93.0    |
| K2255-05 | BFC47            | 5     | 1.19              | 9.59                          | 9.04                              | 93.5    |
| K2255-06 | VHBLK01          | 6     | 1.00              | 2.00                          | 2.00                              | 100.0   |
| K2255-07 | BF8C6            | 7     | 1.15              | 9.71                          | 8.5                               | 85.9    |
| K2255-08 | BF8C6MS          | 8     | 1.15              | 9.71                          | 8.5                               | 85.9    |
| K2255-09 | BF8C6MSD         | 9     | 1.15              | 9.71                          | 8.5                               | 85.9    |
| K2255-10 | BF8C7            | 10    | 1.13              | 9.61                          | 7.82                              | 78.9    |
| K2255-11 | BFC18            | 11    | 1.18              | 9.68                          | 7.18                              | 70.6    |
| K2255-12 | BFC43            | 12    | 1.15              | 9.52                          | 8.17                              | 83.9    |
| K2255-13 | BFC44            | 13    | 1.11              | 9.92                          | 7.92                              | 77.3    |
| K2255-14 | BFC45            | 14    | 1.16              | 9.55                          | 6.98                              | 69.4    |
| K2255-15 | BFC54            | 15    | 1.18              | 9.97                          | 7.93                              | 76.8    |
| K2255-16 | BFC55            | 16    | 1.16              | 9.68                          | 7.44                              | 73.7    |
| K2255-17 | BFC56            | 17    | 1.19              | 9.94                          | 7.38                              | 70.7    |
| K2255-18 | BFC19            | 18    | 1.17              | 9.73                          | 9.11                              | 92.8    |
| K2255-19 | BFC20            | 19    | 1.16              | 9.54                          | 7.89                              | 80.3    |
| K2255-20 | BFC21            | 20    | 1.16              | 9.96                          | 8.12                              | 79.1    |
| K2255-21 | BFC48            | 21    | 1.17              | 9.65                          | 9.1                               | 93.5    |
| K2255-22 | BFC49            | 22    | 1.12              | 9.84                          | 8.81                              | 88.2    |
| K2255-23 | BFC50            | 23    | 1.18              | 9.98                          | 7.81                              | 75.3    |

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$