

LB101885


245.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

41282



A= 0.0000e+000

B= 2.4007e-004

C= -3.6308e-002

Rho= 0.9997216

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.003	-0.003	139	0.000	139				
0.05	0.050	0.040	-0.010	318	0.0 %	318				
0.2	0.200	0.175	-0.025	879	0.0 %	879				
2.5	2.500	2.528	0.028	10683	0.0 %	10683				
5.0	5.000	4.951	-0.049	20776	0.0 %	20776				
7.5	7.500	7.684	0.184	32160	0.0 %	32160				
10.0	10.000	9.874	-0.126	41282	0.0 %	41282				

LB101885 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 04 Apr 2019 16:12:05

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	139	-	0.0000	245.1	PPB	04 Apr 2019 16:13:49	S	Std
0.05 - 1	50.05	318	-	0.0500	245.1	PPB	04 Apr 2019 16:16:07	S	Std
0.2 - 1	50.2	879	-	0.2000	245.1	PPB	04 Apr 2019 16:18:25	S	Std
2.5 - 1	52.5	10683	-	2.5000	245.1	PPB	04 Apr 2019 16:20:42	S	Std
5.0 - 1	55.0	20776	-	5.0000	245.1	PPB	04 Apr 2019 16:22:59	S	Std
7.5 - 1	57.0	32160	-	7.5000	245.1	PPB	04 Apr 2019 16:25:15	S	Std
10.0 - 1	50.0	41282	-	10.0000	245.1	PPB	04 Apr 2019 16:34:37	S	Std
ICV89 - 1	ICV89	16632	3.9565	-245.1		PPB	04 Apr 2019 16:38:16	U	SMPL
ICB89 - 1	ICB89	42	-0.0262	-245.1		PPB	04 Apr 2019 16:40:32	U	SMPL
CCV45 - 1	CCV45	20684	4.9293	-245.1		PPB	04 Apr 2019 16:42:47	U	SMPL
CCB45 - 1	CCB45	44	-0.0257	-245.1		PPB	04 Apr 2019 16:45:02	U	SMPL
CRA - 1	CRA	307	0.0374	-245.1		PPB	04 Apr 2019 16:47:17	U	SMPL
HighStd - 1	HighStd	40986	9.8031	-245.1		PPB	04 Apr 2019 16:49:32	U	SMPL
ChkStd - 1	ChkStd	29158	6.9636	-245.1		PPB	04 Apr 2019 16:51:48	U	SMPL
PB118629BL - 1	PBW	-12	-0.0392	-245.1		PPB	04 Apr 2019 16:54:05	U	SMPL
PB118629BS - 1	LCSW	18653	4.4417	-245.1		PPB	04 Apr 2019 16:56:21	U	SMPL
K2224-01 - 1	033019-LIFT-E-D-B	146	-0.0013	-245.1		PPB	04 Apr 2019 16:58:37	U	SMPL
K2224-01DUP - 1	033019-LIFT-E-D-BDUP	184	0.0079	-245.1		PPB	04 Apr 2019 17:00:54	U	SMPL
K2224-01MS - 1	033019-LIFT-E-D-BMS	18519	4.4095	-245.1		PPB	04 Apr 2019 17:07:01	U	SMPL
K2224-01MSD - 1	033019-LIFT-E-D-BMSD	17552	4.1774	-245.1		PPB	04 Apr 2019 17:09:17	U	SMPL
K2224-02 - 1	033019-LIFT-E-D-M	148	-0.0008	-245.1		PPB	04 Apr 2019 17:11:33	U	SMPL
CCV46 - 1	CCV46	20598	4.9086	-245.1		PPB	04 Apr 2019 17:13:49	U	SMPL
CCB46 - 1	CCB46	41	-0.0265	-245.1		PPB	04 Apr 2019 17:16:04	U	SMPL
K2224-03 - 1	033019-LIFT-E-D-S	205	0.0129	-245.1		PPB	04 Apr 2019 17:18:20	U	SMPL
K2224-04 - 1	033019-LIFT-E-U-B	185	0.0081	-245.1		PPB	04 Apr 2019 17:20:35	U	SMPL
K2224-05 - 1	033019-LIFT-E-U-M	104	-0.0113	-245.1		PPB	04 Apr 2019 17:22:51	U	SMPL
K2224-06 - 1	033019-LIFT-E-U-S	95	-0.0135	-245.1		PPB	04 Apr 2019 17:25:07	U	SMPL
K2242-05 - 1	TWP-02	1116	0.2316	-245.1		PPB	04 Apr 2019 17:27:23	U	SMPL
K2224-01LX5 - 1		93	-0.0140	-245.1		PPB	04 Apr 2019 17:29:39	U	SMPL
K2224-01A - 1		17806	4.2384	-245.1		PPB	04 Apr 2019 17:36:01	U	SMPL
CCV47 - 1	CCV47	20427	4.8676	-245.1		PPB	04 Apr 2019 17:38:17	U	SMPL
CCB47 - 1	CCB47	26	-0.0301	-245.1		PPB	04 Apr 2019 17:40:34	U	SMPL