


LB101896

7471B

INSTRUMENT ID: C01

Linear



A= 0.0000e+000

B= 1.7435e-004

C= -3.6281e-003

Rho= 0.9987516

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.029	0.029	189	0.000	189				
0.2	0.200	0.195	-0.005	1137	0.0 %	1137				
2.5	2.500	2.614	0.114	15015	0.0 %	15015				
5.0	5.000	4.681	-0.319	26870	0.0 %	26870				
7.5	7.500	7.780	0.280	44643	0.0 %	44643				
10.0	10.000	9.901	-0.099	56805	0.0 %	56805				

LB101896 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 05 Apr 2019 11:58:27

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	189	-	0.0000	7471B	PPB	05 Apr 2019 12:31:56	S	Std
0.2 - 1	50.2	1137	-	0.2000	7471B	PPB	05 Apr 2019 12:34:15	S	Std
2.5 - 1	52.5	15015	-	2.5000	7471B	PPB	05 Apr 2019 12:36:32	S	Std
5.0 - 1	55.0	26870	-	5.0000	7471B	PPB	05 Apr 2019 12:38:49	S	Std
7.5 - 1	57.5	44643	-	7.5000	7471B	PPB	05 Apr 2019 12:41:06	S	Std
10.0 - 1	510.0	56805	-	10.0000	7471B	PPB	05 Apr 2019 12:46:34	S	Std
ICV90 - 1	ICV90	24705	4.3038	-	7471B	PPB	05 Apr 2019 12:50:11	U	SMPL
ICB90 - 1	ICB90	106	0.0149	-	7471B	PPB	05 Apr 2019 12:52:27	U	SMPL
CCV48 - 1	CCV48	26882	4.6833	-	7471B	PPB	05 Apr 2019 12:54:43	U	SMPL
CCB48 - 1	CCB48	30	0.0016	-	7471B	PPB	05 Apr 2019 12:56:58	U	SMPL
CRA - 1	CRA	1106	0.1892	-	7471B	PPB	05 Apr 2019 12:59:14	U	SMPL
HighStd - 1	HighStd	49488	8.6248	-	7471B	PPB	05 Apr 2019 13:01:30	U	SMPL
ChkStd - 1	ChkStd	40574	7.0706	-	7471B	PPB	05 Apr 2019 13:03:45	U	SMPL
PB118652BL - 1	PBS	1	-0.0035	-	7471B	PPB	05 Apr 2019 13:06:04	U	SMPL
PB118652BS - 1	LCSS	23317	4.0618	-	7471B	PPB	05 Apr 2019 13:08:21	U	SMPL
K2264-01 - 1	OPEN-WALL-EXCAVATION	12505	2.1767	-	7471B	PPB	05 Apr 2019 13:10:37	U	SMPL
K2264-01DUP - 1	OPEN-WALL-EXCAVATIONDUP	12482	2.1726	-	7471B	PPB	05 Apr 2019 13:12:54	U	SMPL
K2264-01MS - 1	OPEN-WALL-EXCAVATIONMS	36973	6.4427	-	7471B	PPB	05 Apr 2019 13:15:11	U	SMPL
K2264-01MSD - 1	OPEN-WALL-EXCAVATIONMSD	35325	6.1554	-	7471B	PPB	05 Apr 2019 13:17:28	U	SMPL
K2264-04 - 1	293-72-11966-238	61292	10.6828	-	7471B	PPB	05 Apr 2019 13:19:45	U	SMPL
CCV49 - 1	CCV49	29482	5.1367	-	7471B	PPB	05 Apr 2019 13:22:01	U	SMPL
CCB49 - 1	CCB49	-49	-0.0122	-	7471B	PPB	05 Apr 2019 13:24:18	U	SMPL
K2264-01LX5 - 1		2189	0.3780	-	7471B	PPB	05 Apr 2019 13:26:35	U	SMPL
K2264-01A - 1		35953	6.2649	-	7471B	PPB	05 Apr 2019 13:28:50	U	SMPL
K2264-04DLX2 - 1		30755	5.3586	-	7471B	PPB	05 Apr 2019 13:31:05	U	SMPL
CCV50 - 1	CCV50	28666	4.9944	-	7471B	PPB	05 Apr 2019 13:33:22	U	SMPL
CCB50 - 1	CCB50	-41	-0.0108	-	7471B	PPB	05 Apr 2019 13:35:39	U	SMPL