



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB101899

Ambient Barometric Pressure (mmHg): 771

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.0	80.6	79.9	04/05/2019	09:45
2	K2265-02		1	8	>100.0	>212.0	>212.0	04/05/2019	10:20
3	K2265-02DUP		1	8	>100.0	>212.0	>212.0	04/05/2019	10:50

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB101899

WorkList Name : FP4519

WorkList ID : 12447

Date : 04-05-2019 09:17:04

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/11/2019	Water	K2265-02	C	Cool 4 deg C	PSEG01	K31	BW-04419	04/04/2019	1010A

Date/Time 04/05/19 09:20
 Received by: KS
 Relinquished by: CP

Date/Time 04/05/19 10:55
 Received by: CP
 Relinquished by: KS