



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 4/8/2019

OVENTEMP IN Celsius(°C): 108
Time IN: 16:05
In Date: 04/05/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:30
Out Date: 04/06/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-1
Thermometer ID: %SOLIDS-OVEN

QC:LB101909

Lab ID	Client Sample ID	Dish#	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish + Dry Sample Wt(g) (C)	% Solid
K2303-01	BF896	1	1.15	9.75	8.67	87.4
K2303-02	BF897	2	1.13	9.68	7.82	78.2
K2303-03	BF898	3	1.14	9.55	8.74	90.4
K2303-04	BF899	4	1.17	9.6	8.65	88.7
K2303-05	BF8A0	5	1.18	9.96	7.79	75.3
K2303-06	BF8A1	6	1.14	9.97	7.96	77.2
K2303-07	BF8A2	7	1.17	9.6	7.41	74.0
K2303-08	BF8A3	8	1.16	9.66	7.21	71.2
K2303-09	BF8A4	9	1.13	9.9	8.19	80.5
K2303-10	BF8B1	10	1.12	9.69	8.04	80.7
K2303-11	BF881	11	1.18	9.77	7.52	73.8
K2303-12	BF881MS	12	1.18	9.77	7.52	73.8
K2303-13	BF881MSD	13	1.18	9.77	7.52	73.8
K2303-14	BF882	14	1.18	9.98	8.04	78.0
K2303-15	BF883	15	1.13	9.65	7.71	77.2
K2303-16	BF884	16	1.15	9.95	8.38	82.2
K2303-17	BF885	17	1.17	9.92	7.89	76.8
K2303-18	BF886	18	1.14	9.94	7.95	77.4
K2303-19	BF887	19	1.17	9.55	7.36	73.9
K2303-20	BF8B3	20	1.12	9.59	7.28	72.7
K2303-24	VHBLK02	21	1.00	2.00	2.00	100.0

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$