



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB101916

Slope : 98.5

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	04/05/2019	17:00
2	CAL2	1	Water	NA	NA	20.1	7.01	04/05/2019	17:01
3	CAL3	1	Water	NA	NA	20.1	10.02	04/05/2019	17:02
4	ICV	1	Water	NA	NA	20.1	7.00	04/05/2019	17:05
5	CCV1	1	Water	NA	NA	20.1	2.01	04/05/2019	17:06
6	K2294-04	1	Solid	20.03	20	21.6	7.19	04/05/2019	17:08
7	K2294-04DUP	1	Solid	20.04	20	21.6	7.20	04/05/2019	17:10
8	K2294-08	1	Solid	20.02	20	21.2	6.02	04/05/2019	17:11
9	K2309-02	1	Solid	20.03	20	21.2	7.11	04/05/2019	17:12
10	CCV2	1	Water	NA	NA	20.2	12.02	04/05/2019	17:15

WORKLIST(Hardcopy Internal Chain)

LB 101916

WorkList Name : CORROSIVITY K2294

WorkList ID : 124479

Date : 04-05-2019 15:42:18

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/05/2019	Solid	K2294-04	Corrosivity	Cool 4 deg C	PSEG01	K31	TP-14-S	04/05/2019	9045D
04/05/2019	Solid	K2294-08	Corrosivity	Cool 4 deg C	PSEG01	K31	TP-14-D	04/05/2019	9045D
04/05/2019	Solid	K2309-02	Corrosivity	Cool 4 deg C	PSEG01	K31	HA-040519	04/05/2019	9045D

Date/Time 04-05-19 16:00
 Received by: J1
 Relinquished by: CR

Date/Time 04-05-19 17:30
 Received by: CR
 Relinquished by: Jh