



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 4/11/2019

OVENTEMP IN Celsius(°C): 107
Time IN: 15:30
In Date: 04/08/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:30
Out Date: 04/09/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-1
Thermometer ID: %SOLIDS-OVEN

QC:LB101924

Lab ID	Client Sample ID	Dish#	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish + Dry Sample Wt(g) (C)	% Solid
K2278-01	DB6W8	1	1.00	2.00	2.00	100.0
K2278-02	DB6W9	2	1.13	9.85	8.44	83.8
K2278-03	DB6X0	3	1.16	9.95	8.47	83.2
K2278-04	DB6X1	4	1.12	9.57	7.93	80.6
K2278-05	DB6X2	5	1.12	9.55	8.5	87.5
K2278-06	DB6X3	6	1.15	9.8	8.93	89.9
K2278-07	DB6X4	7	1.17	9.73	8.61	86.9
K2278-08	DB6X5	8	1.18	9.97	8.00	77.6
K2278-09	VHBLK01	9	1.00	2.00	2.00	100.0
K2278-10	DB6Y0	10	1.13	9.79	9.24	93.6
K2278-11	DB6Y2	11	1.14	9.58	8.18	83.4
K2278-12	DB6Y3	12	1.16	9.8	8.55	85.5
K2278-13	DB6Y3MS	13	1.16	9.8	8.55	85.5
K2278-14	DB6Y3MSD	14	1.16	9.8	8.55	85.5
K2278-15	DB6Y7	15	1.18	9.6	7.97	80.6
K2278-16	DB6Y8	16	1.13	9.71	8.00	80.1
K2278-17	DB6Y9	17	1.14	9.81	8.39	83.6
K2278-18	DB6Z0	18	1.14	9.98	8.41	82.2
K2278-19	DB6Z1	19	1.18	9.81	8.57	85.6
K2278-20	DB6Z2	20	1.1	9.62	8.3	84.5
K2278-21	DB6Z4	21	1.19	9.73	8.37	84.1
K2278-22	DB6Z5	22	1.13	9.73	8.05	80.5
K2278-23	DB6Z6	23	1.13	9.77	8.22	82.1
K2278-24	DB701	24	1.18	9.97	7.84	75.8

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$