

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

Instrument ID : Konelab

4/8/2019 14:54

Test: Phenolics-

mg/d

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.024	0.0	0.122	
ICB1	-0.005	0.0	0.012	
CCV1	0.987	0.0	0.118	
CCB1	-0.004	0.0	0.012	
PB118662BL	-0.001	0.0	0.013	
PB118662BS	1.012	0.0	0.121	
K2242-05	-0.001	0.0	0.013	
K2242-05DUP	-0.002	0.0	0.013	
K2242-05MS	0.904	0.0	0.109	
K2242-05MSD	0.903	0.0	0.109	
K2256-02	0.004	0.0	0.013	
K2241-11	0.097	0.0	0.023	
K2241-10	0.006	0.0	0.013	
CCV2	0.992	0.0	0.119	
CCB2	-0.001	0.0	0.013	

N 15
Mean 0.394
SD 0.4885
CV% 123.90

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AB

Instrument ID : Konelab

4/8/2019 14:15


Test Phenolics-

Accepted 4/8/2019 14:15

Factor 9.369
Bias 0.013

Coeff. of det. 0.999951

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.018	0.0524	0.0500	
2	Phenol-2	0.024	0.1030	0.1000	
3	Phenol-2	0.040	0.2527	0.2500	
4	Phenol-2	0.066	0.4957	0.5000	
5	Phenol-2	0.119	0.9914	1.0000	
6	Phenol-2	0.227	2.0048	2.0000	

AS

Aquakem v. 7.2AQ1

Results from time period:

Mon Apr 08 14:15:03 2019

Mon Apr 08 14:47:15 2019

Sample Id	Sam/Ctr/c/	Test short nan	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0524	mg/l	4/8/2019 14:15:03	
0.1PPM	A	Phenolics-	P	0.103	mg/l	4/8/2019 14:15:04	
0.25PPM	A	Phenolics-	P	0.2527	mg/l	4/8/2019 14:15:05	
0.50PPM	A	Phenolics-	P	0.4957	mg/l	4/8/2019 14:15:06	
1.0PPM	A	Phenolics-	P	0.9914	mg/l	4/8/2019 14:15:07	
2.0PPM	A	Phenolics-	P	2.0048	mg/l	4/8/2019 14:15:08	
ICV1	S	Phenolics-	P	1.0242	mg/l	4/8/2019 14:40:51	
ICB1	S	Phenolics-	P	-0.005	mg/l	4/8/2019 14:40:52	
CCV1	S	Phenolics-	P	0.9869	mg/l	4/8/2019 14:40:53	
CCB1	S	Phenolics-	P	-0.004	mg/l	4/8/2019 14:40:54	
PB118662BL	S	Phenolics-	P	-0.0006	mg/l	4/8/2019 14:40:55	
PB118662BS	S	Phenolics-	P	1.0116	mg/l	4/8/2019 14:40:56	
K2242-05	S	Phenolics-	P	-0.0013	mg/l	4/8/2019 14:40:57	
K2242-05DUP	S	Phenolics-	P	-0.0019	mg/l	4/8/2019 14:40:58	
K2242-05MS	S	Phenolics-	P	0.904	mg/l	4/8/2019 14:40:59	
K2242-05MSD	S	Phenolics-	P	0.9032	mg/l	4/8/2019 14:41:00	
K2256-02	S	Phenolics-	P	0.0037	mg/l	4/8/2019 14:41:01	
K2241-11	S	Phenolics-	P	0.0969	mg/l	4/8/2019 14:47:10	
K2241-10	S	Phenolics-	P	0.0059	mg/l	4/8/2019 14:47:11	
CCV2	S	Phenolics-	P	0.9919	mg/l	4/8/2019 14:47:12	
CCB2	S	Phenolics-	P	-0.0012	mg/l	4/8/2019 14:47:13	