

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB101940

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP70044
calibration std. phosphate 1 ppm	WP73402
calibration std. phosphate 0.5 ppm	WP73401
calibration std. phosphate 0.3 ppm	WP73400
calibration std. 0 ppm	WP73397
phosphate ICV std.	WP73404
phosphate CCV std.	WP73403
calibration std. phosphate 0.1 ppm	WP73399
calibration std. phosphate 0.05 ppm	WP73398
Combined reagent	WP73407
11 N sulfuric acid	WP71598
Phenolphthalein indicator	WP71599
Sodium hydroxide, 1N	WP71600
Anions 300/9056 calibration standard 5-CCV	WP71405
Anions 300/9056 calibration standard 4	WP71406

Intercept: -0.0008Slope: 0.6522Regression: 0.999811

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	04/08/2019	14:40
2	CAL2	0.05	1	50	50	0.033	04/08/2019	14:42
3	CAL3	0.10	1	50	50	0.066	04/08/2019	14:42
4	CAL4	0.30	1	50	50	0.186	04/08/2019	14:43
5	CAL5	0.50	1	50	50	0.331	04/08/2019	14:43
6	CAL6	1.00	1	50	50	0.651	04/08/2019	14:44



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB101940

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.318	0.489	04/08/2019	14:45
2	ICB		1	50	50	0.000	0.001	04/08/2019	14:46
3	CCV1	0.50	1	50	50	0.328	0.504	04/08/2019	14:50
4	CCB1		1	50	50	0.000	0.001	04/08/2019	14:50
5	PB118696BL		1	50	50	0.000	0.001	04/08/2019	14:51
6	PB118696BS	0.50	1	50	50	0.310	0.477	04/08/2019	14:52
7	K2241-07		1	50	50	0.015	0.024	04/08/2019	14:53
8	K2241-08		1	50	50	0.061	0.095	04/08/2019	14:54
9	K2307-01		1	50	50	3.603	5.526	04/08/2019	14:55
10	K2331-03		1	50	50	0.068	0.105	04/08/2019	14:56
11	K2332-02		1	50	50	0.244	0.375	04/08/2019	14:56
12	K2333-03		1	50	50	0.332	0.510	04/08/2019	14:57
13	K2333-03DUP		1	50	50	0.330	0.507	04/08/2019	14:58
14	K2333-03MS	0.50	1	50	50	0.612	0.940	04/08/2019	14:58
15	CCV2	0.50	1	50	50	0.324	0.498	04/08/2019	14:59
16	CCB2		1	50	50	0.000	0.001	04/08/2019	15:00
17	K2333-03MSD	0.50	1	50	50	0.620	0.952	04/08/2019	15:01
18	K2307-01		50	50	50	0.648	0.995	04/08/2019	15:15
19	CCV3	0.50	1	50	50	0.326	0.501	04/08/2019	15:15
20	CCB3		1	50	50	0.000	0.001	04/08/2019	15:16