



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB101947

Slope : 98.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	04/08/2019	16:30
2	CAL2	1	Water	NA	NA	20.1	7.01	04/08/2019	16:31
3	CAL3	1	Water	NA	NA	20.1	10.02	04/08/2019	16:32
4	ICV	1	Water	NA	NA	20.2	7.00	04/08/2019	16:33
5	CCV1	1	Water	NA	NA	20.2	2.02	04/08/2019	16:35
6	K2337-01	1	Water	NA	NA	19.5	5.73	04/08/2019	16:36
7	K2338-05	1	Water	NA	NA	21.6	6.51	04/08/2019	16:37
8	K2338-05DUP	1	Water	NA	NA	21.7	6.52	04/08/2019	16:38
9	CCV2	1	Water	NA	NA	20.2	12.02	04/08/2019	16:40

WORKLIST(Hardcopy Internal Chain)

VB101947

WorkList Name : PH K2195

WorkList ID : 124537

Date : 04-08-2019 16:13:22

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/08/2019	Water	K2337-01-9	pH	Cool 4 deg C	PSEG01	K31	WS040819	04/08/2019	9040C
04/08/2019	Water	K2338-05 -17	pH	Cool 4 deg C	PSEG01	K31	TP-1-WATER	04/08/2019	9040C

Date/Time

Received by:

Relinquished by:

04-08-19 16:15

JP

BT

Date/Time

Received by:

Relinquished by:

04-08-19 18:35

BT

SN