



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB101948

Slope : 98.5

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	04/08/2019	16:30
2	CAL2	1	Water	NA	NA	20.1	7.01	04/08/2019	16:31
3	CAL3	1	Water	NA	NA	20.1	10.02	04/08/2019	16:32
4	ICV	1	Water	NA	NA	20.2	7.00	04/08/2019	16:33
5	CCV1	1	Water	NA	NA	20.2	2.01	04/08/2019	16:50
6	K2198-02	1	Solid	20.02	20	23.8	7.41	04/08/2019	16:51
7	K2199-06	1	Solid	20.03	20	23.7	7.99	04/08/2019	17:00
8	K2199-08	1	Solid	20.01	20	23.8	7.31	04/08/2019	17:01
9	K2337-04	1	Solid	20.03	20	22.8	5.58	04/08/2019	17:03
10	K2338-02	1	Solid	20.04	20	23.8	7.41	04/08/2019	17:05
11	K2338-04	1	Solid	20.02	20	23.7	6.95	04/08/2019	17:06
12	K2338-09	1	Solid	20.04	20	23.6	7.52	04/08/2019	17:08
13	K2338-11	1	Solid	20.03	20	23.6	7.37	04/08/2019	17:10
14	K2338-13	1	Solid	20.05	20	23.6	7.67	04/08/2019	17:11
15	K2338-13DUP	1	Solid	20.04	20	23.7	7.67	04/08/2019	17:13
16	CCV2	1	Water	NA	NA	20.1	12.01	04/08/2019	17:15

15101948

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY K2199

WorkList ID : 124535

Date : 04-08-2019 16:11:58

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/08/2019	Solid	K2198-02	Corrosivity	Cool 4 deg C	PSEG05	K31	OK-01-040819	04/08/2019	9045D
04/08/2019	Solid	K2199-06	Corrosivity	Cool 4 deg C	PSEG05		OR-03-040819-A	04/08/2019	9045D
04/08/2019	Solid	K2199-08	Corrosivity	Cool 4 deg C	PSEG05		OR-03-040819-B	04/08/2019	9045D
04/08/2019	Solid	K2337-04	Corrosivity	Cool 4 deg C	PSEG01	K31	ODS-040819	04/08/2019	9045D
04/08/2019	Solid	K2338-02	Corrosivity	Cool 4 deg C	PSEG01	K31	TP-1-SHALLOW	04/08/2019	9045D
04/08/2019	Solid	K2338-04	Corrosivity	Cool 4 deg C	PSEG01	K31	TP-1-DEEP	04/08/2019	9045D
04/08/2019	Solid	K2338-09	Corrosivity	Cool 4 deg C	PSEG01	K31	TP-2-SHALLOW	04/08/2019	9045D
04/08/2019	Solid	K2338-11	Corrosivity	Cool 4 deg C	PSEG01	K31	TP-2-DEEP	04/08/2019	9045D
04/08/2019	Solid	K2338-13	Corrosivity	Cool 4 deg C	PSEG01	K31	TP-3-SHALLOW	04/08/2019	9045D

Date/Time 04.08.19 16.15
 Received by: J.
 Relinquished by: A

Date/Time 04.08.19 18.35
 Received by: A
 Relinquished by: J.