

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus, TotalSUPERVISOR REVIEW BY: apatelRun Number: LB101965BALANCE ID: WC SC-4

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP70044
calibration std. phosphate 1 ppm	WP73402
calibration std. phosphate 0.5 ppm	WP73401
calibration std. phosphate 0.3 ppm	WP73400
calibration std. 0 ppm	WP73397
phosphate ICV std.	WP73404
phosphate CCV std.	WP73403
calibration std. phosphate 0.1 ppm	WP73399
calibration std. phosphate 0.05 ppm	WP73398
Combined reagent	WP73444
11 N sulfuric acid	WP71598
Phenolphthalein indicator	WP71599
Sodium hydroxide, 1N	WP71600
PHOSPATE LOD STD, 0.025PPM	WP73445
PHOSPATE LOQ STD, 0.1PPM	WP73449

Intercept: -0.0007Slope: 0.6513Regression: 0.999809

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	04/09/2019	12:53
2	CAL2	0.05	1	50	50	0.033	04/09/2019	12:56
3	CAL3	0.10	1	50	50	0.066	04/09/2019	12:56
4	CAL4	0.30	1	50	50	0.186	04/09/2019	12:57
5	CAL5	0.50	1	50	50	0.331	04/09/2019	12:58
6	CAL6	1.00	1	50	50	0.650	04/09/2019	12:59



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus, Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB101965

BALANCE ID: WC SC-4

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.315	0.485	04/09/2019	13:00
2	ICB		1	50	50	0.000	0.001	04/09/2019	13:00
3	CCV1	0.50	1	50	50	0.326	0.502	04/09/2019	13:08
4	CCB1		1	50	50	0.000	0.001	04/09/2019	13:08
5	PB118737BL		1	1.00	50	0.000	0.001	04/09/2019	13:09
6	PB118737BS	0.50	1	1.01	50	0.306	0.471	04/09/2019	13:10
7	K2241-01		1	1.01	50	0.013	0.021	04/09/2019	13:10
8	K2241-02		1	1.02	50	0.060	0.093	04/09/2019	13:11
9	CCV2	0.50	1	50	50	0.327	0.503	04/09/2019	13:11
10	CCB2		1	50	50	0.000	0.001	04/09/2019	13:12