



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB101972

Slope : 98.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	04/09/2019	15:30
2	CAL2	1	Water	NA	NA	20.1	7.01	04/09/2019	15:31
3	CAL3	1	Water	NA	NA	20.1	10.02	04/09/2019	15:33
4	ICV	1	Water	NA	NA	20.1	7.00	04/09/2019	15:35
5	CCV1	1	Water	NA	NA	20.1	2.01	04/09/2019	15:36
6	K2335-02	1	Water	NA	NA	20.8	7.18	04/09/2019	15:37
7	K2335-02DUP	1	Water	NA	NA	20.9	7.19	04/09/2019	15:38
8	K2342-18	1	Water	NA	NA	11.8	6.93	04/09/2019	15:40
9	CCV2	1	Water	NA	NA	20.2	12.01	04/09/2019	15:41

WORKLIST(Hardcopy Internal Chain)

28101972

WorkList Name : CORROSIONITY K2335

WorkList ID : 124594

Date : 04-09-2019 14:51:21

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/09/2019	Water	K2335-02-1A	Corrosivity	Cool 4 deg C	PSEG01	K31	SV30755L	04/09/2019	9040C
04/08/2019	Water	K2342-18 -1X	Corrosivity	Cool 4 deg C	PSEG01	K31	TP-12	04/08/2019	9040C

Date/Time 04-09-19 15:00

Received by: JD

Relinquished by: PR

Date/Time 04-09-19 16:00

Received by: PR

Relinquished by: JH