



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB101974

Slope : 98.5

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	04/09/2019	17:10
2	CAL2	1	Water	NA	NA	20.2	7.01	04/09/2019	17:11
3	CAL3	1	Water	NA	NA	20.2	10.03	04/09/2019	17:12
4	ICV	1	Water	NA	NA	20.2	7.00	04/09/2019	17:15
5	CCV1	1	Water	NA	NA	20.1	2.01	04/09/2019	17:16
6	K2200-02	1	Solid	20.03	20	22.8	9.25	04/09/2019	17:18
7	K2200-02DUP	1	Solid	20.04	20	22.9	9.26	04/09/2019	17:19
8	K2200-04	1	Solid	20.02	20	22.5	8.94	04/09/2019	17:20
9	K2200-06	1	Solid	20.04	20	21.7	9.20	04/09/2019	17:21
10	K2203-06	1	Solid	20.02	20	22.4	8.16	04/09/2019	17:22
11	K2203-08	1	Solid	20.01	20	22.0	8.66	04/09/2019	17:24
12	K2334-03	1	Solid	20.03	20	22.6	8.82	04/09/2019	17:25
13	K2342-04	1	Solid	20.05	20	22.6	8.41	04/09/2019	17:26
14	K2342-08	1	Solid	20.04	20	22.3	7.45	04/09/2019	17:29
15	K2342-12	1	Solid	20.02	20	22.4	8.68	04/09/2019	17:30
16	CCV2	1	Water	NA	NA	20.1	12.01	04/09/2019	17:31
17	K2342-16	1	Solid	20.03	20	21.9	7.31	04/09/2019	17:33
18	K2343-04	1	Solid	20.02	20	21.6	8.54	04/09/2019	17:35
19	K2343-08	1	Solid	20.03	20	21.7	8.39	04/09/2019	17:36
20	K2344-03	1	Solid	20.01	20	22.6	5.56	04/09/2019	17:39

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	K2344-05	1	Solid	20.00	20	22.5	5.58	04/09/2019	17:40
22	K2346-03	1	Solid	20.03	20	22.2	8.54	04/09/2019	17:41
23	CCV3	1	Water	NA	NA	20.1	2.02	04/09/2019	17:43

WORKLIST(Hardcopy Internal Chain)

13101944

WorkList Name : CORROSIONITY K2200

WorkList ID : 124592

Date : 04-09-2019 14:49:49

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/09/2019	Solid	K2200-02	Corrosivity	Cool 4 deg C	PSEG05	K31	EO-01-040919-A	04/09/2019	9045D
04/09/2019	Solid	K2200-04	Corrosivity	Cool 4 deg C	PSEG05	K31	EO-01-040919-B	04/09/2019	9045D
04/09/2019	Solid	K2200-06	Corrosivity	Cool 4 deg C	PSEG05	K31	EO-01-040919-C	04/09/2019	9045D
04/09/2019	Solid	K2203-06	Corrosivity	Cool 4 deg C	PSEG05		SU-04-040919-A	04/09/2019	9045D
04/09/2019	Solid	K2203-08	Corrosivity	Cool 4 deg C	PSEG05		SU-04-040919-B	04/09/2019	9045D
04/08/2019	Solid	K2334-03	Corrosivity	Cool 4 deg C	WSP001	F11	P-6(0-10)	04/08/2019	9045D
04/09/2019	Solid	K2342-04	Corrosivity	Cool 4 deg C	PSEG01	K31	TP-11-S	04/09/2019	9045D
04/09/2019	Solid	K2342-08	Corrosivity	Cool 4 deg C	PSEG01	K31	TP-11-D	04/09/2019	9045D
04/08/2019	Solid	K2342-12	Corrosivity	Cool 4 deg C	PSEG01	K31	TP-12-S	04/08/2019	9045D
04/08/2019	Solid	K2342-16	Corrosivity	Cool 4 deg C	PSEG01	K31	TP-12-D	04/08/2019	9045D
04/08/2019	Solid	K2343-04	Corrosivity	Cool 4 deg C	PSEG01	K31	TP-7-S	04/08/2019	9045D
04/08/2019	Solid	K2343-08	Corrosivity	Cool 4 deg C	PSEG01	K31	TP-7-D	04/08/2019	9045D
04/09/2019	Solid	K2344-03	Corrosivity	Cool 4 deg C	PSEG01	K31	WOOD-1	04/09/2019	9045D
04/09/2019	Solid	K2344-05	Corrosivity	Cool 4 deg C	PSEG01	K31	WOOD-2	04/09/2019	9045D
04/09/2019	Solid	K2346-03	Corrosivity	Cool 4 deg C	PSEG01	K31	VNJ-229	04/09/2019	9045D

Date/Time 04-09-19 16:15
 Received by: Jo
 Relinquished by: Jo

Date/Time 04-09-19 18:30
 Received by: Jo
 Relinquished by: Jo