



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB101975

Ambient Barometric Pressure (mmHg): 760

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.0	80.6	80.6	04/09/2019	10:05
2	K2337-02		1	8	>100.0	>212.0	>212.0	04/09/2019	10:35
3	K2337-02DUP		1	8	>100.0	>212.0	>212.0	04/09/2019	10:55
4	K2338-06		1	9	>100.0	>212.0	>212.0	04/09/2019	11:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 101975

WorkList Name : FP4919

WorkList ID : 124575

Date : 04-09-2019 09:38:58

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/12/2019	Water	K2337-02 C	Flash Point	Cool 4 deg C	PSEG01	K31	WS040819	04/08/2019	1010A
04/22/2019	Water	K2338-06 C	Flash Point	Cool 4 deg C	PSEG01	K31	TP-1-WATER	04/08/2019	1010A

Date/Time 04/09/19 09:40
 Received by: KS
 Relinquished by: CP

Date/Time 04/09/19 11:25
 Received by: CP
 Relinquished by: KS