

LB101982


7471B

INSTRUMENT ID : CU1

Linear

 μ Abs.:

59495



A= 0.0000e+000

B= 1.6847e-004

C= -8.4218e-002

Rho= 0.9994925

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.020	-0.020	379	0.000	379				
0.2	0.200	0.167	-0.033	1494	0.0 %	1494				
2.5	2.500	2.435	-0.065	14951	0.0 %	14951				
5.0	5.000	5.259	0.259	31716	0.0 %	31716				
7.5	7.500	7.420	-0.080	44545	0.0 %	44545				
10.0	10.000	9.939	-0.061	59495	0.0 %	59495				

LB101982 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 10 Apr 2019 09:06:18

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50-0	379	-	0.00007471B	PPB		10 Apr 2019 09:10:00	S	Std
0.2 - 1	50-2	1494	-	0.20007471B	PPB		10 Apr 2019 09:12:18	S	Std
2.5 - 1	52-5	14951	-	2.50007471B	PPB		10 Apr 2019 09:14:35	S	Std
5.0 - 1	55-0	31716	-	5.00007471B	PPB		10 Apr 2019 09:16:52	S	Std
7.5 - 1	57-5	44545	-	7.50007471B	PPB		10 Apr 2019 09:19:08	S	Std
10.0 - 1	510-0	59495	-	10.00007471B	PPB		10 Apr 2019 09:21:25	S	Std
ICV92 - 1	ICV92	23848	3.9335	-7471B	PPB		10 Apr 2019 09:25:14	U	SMPL
ICB92 - 1	ICB92	322	-0.0300	-7471B	PPB		10 Apr 2019 09:27:29	U	SMPL
CCV58 - 1	CCV58	31910	5.2917	-7471B	PPB		10 Apr 2019 09:29:44	U	SMPL
CCB58 - 1	CCB58	319	-0.0305	-7471B	PPB		10 Apr 2019 09:31:59	U	SMPL
CRA - 1	CRA	1527	0.1730	-7471B	PPB		10 Apr 2019 09:34:15	U	SMPL
HighStd - 1	HighStd	60021	10.0276	-7471B	PPB		10 Apr 2019 09:36:30	U	SMPL
ChkStd - 1	ChkStd	41908	6.9761	-7471B	PPB		10 Apr 2019 09:38:45	U	SMPL
PB118745BL - 1	PBS	-12	-0.0862	-7471B	PPB		10 Apr 2019 09:41:03	U	SMPL
PB118745BS - 1	LCSS	21708	3.5730	-7471B	PPB		10 Apr 2019 09:43:21	U	SMPL
K2198-01 - 1	OK-01-040819	1642	0.1924	-7471B	PPB		10 Apr 2019 09:45:37	U	SMPL
K2198-01DUP - 1	OK-01-040819DUP	1627	0.1899	-7471B	PPB		10 Apr 2019 09:47:54	U	SMPL
K2198-01MS - 1	OK-01-040819MS	23958	3.9520	-7471B	PPB		10 Apr 2019 09:50:11	U	SMPL
K2198-01MSD - 1	OK-01-040819MSD	23929	3.9471	-7471B	PPB		10 Apr 2019 09:52:28	U	SMPL
K2199-05 - 1	OR-03-040819-A	3911	0.5747	-7471B	PPB		10 Apr 2019 09:54:44	U	SMPL
CCV59 - 1	CCV59	30305	5.0213	-7471B	PPB		10 Apr 2019 09:56:58	U	SMPL
CCB59 - 1	CCB59	274	-0.0381	-7471B	PPB		10 Apr 2019 09:59:14	U	SMPL
K2199-07 - 1	OR-03-040819-B	3210	0.4566	-7471B	PPB		10 Apr 2019 10:01:30	U	SMPL
K2294-03 - 1	TP-14-S	8571	1.3597	-7471B	PPB		10 Apr 2019 10:03:45	U	SMPL
K2294-07 - 1	TP-14-D	5835	0.8988	-7471B	PPB		10 Apr 2019 10:06:01	U	SMPL
K2298-01 - 1	WALL-PAINT	44	-0.0768	-7471B	PPB		10 Apr 2019 10:08:17	U	SMPL
K2298-03 - 1	FLOOR-PAINT	10	-0.0825	-7471B	PPB		10 Apr 2019 10:10:33	U	SMPL
K2298-05 - 1	CELLING-PAINT	83	-0.0702	-7471B	PPB		10 Apr 2019 10:12:49	U	SMPL
K2305-01 - 1	FK-EF-02-001-ABCD	2411	0.3220	-7471B	PPB		10 Apr 2019 10:15:06	U	SMPL
K2309-01 - 1	HA-040519	13142	2.1298	-7471B	PPB		10 Apr 2019 10:17:23	U	SMPL
K2316-01 - 1	20190169-COMP-CS-1	105545	17.6971	-7471B	PPB		10 Apr 2019 10:19:40	U	SMPL
K2316-02 - 1	20190170-COMP-CS-2	74421	12.4536	-7471B	PPB		10 Apr 2019 10:21:56	U	SMPL
CCV60 - 1	CCV60	32135	5.3296	-7471B	PPB		10 Apr 2019 10:24:14	U	SMPL
CCB60 - 1	CCB60	12	-0.0822	-7471B	PPB		10 Apr 2019 10:26:32	U	SMPL
K2316-03 - 1	20190171-COMP-CS-3	104580	17.5345	-7471B	PPB		10 Apr 2019 10:28:48	U	SMPL
K2316-04 - 1	20190172-COMP-CS-4	94066	15.7632	-7471B	PPB		10 Apr 2019 10:31:04	U	SMPL
K2316-05 - 1	20190173-COMP-CS-5	63045	10.5370	-7471B	PPB		10 Apr 2019 10:33:21	U	SMPL
K2316-06 - 1	20190174-AMND-COMP-CS-1	74936	12.5403	-7471B	PPB		10 Apr 2019 10:35:39	U	SMPL
K2316-07 - 1	20190175-AMND-COMP-CS-2	62791	10.4942	-7471B	PPB		10 Apr 2019 10:37:56	U	SMPL
K2316-08 - 1	20190176-AMND-COMP-CS-3	96796	16.2231	-7471B	PPB		10 Apr 2019 10:40:15	U	SMPL
K2316-09 - 1	20190177-AMND-COMP-CS-4	75761	12.6793	-7471B	PPB		10 Apr 2019 10:42:32	U	SMPL
K2316-10 - 1	20190178-AMND-COMP-CS-5	45324	7.5516	-7471B	PPB		10 Apr 2019 10:44:51	U	SMPL
PB118746BL - 1	PBS	-35	-0.0901	-7471B	PPB		10 Apr 2019 10:47:10	U	SMPL
PB118746BS - 1	LCSS	22340	3.6794	-7471B	PPB		10 Apr 2019 10:49:27	U	SMPL
CCV61 - 1	CCV61	30854	5.1138	-7471B	PPB		10 Apr 2019 10:51:43	U	SMPL
CCB61 - 1	CCB61	216	-0.0478	-7471B	PPB		10 Apr 2019 10:53:59	U	SMPL
K2317-01 - 1	SB-1_4.5-5.0	510	0.0017	-7471B	PPB		10 Apr 2019 10:56:15	U	SMPL
K2317-01DUP - 1	SB-1_4.5-5.0DUP	508	0.0014	-7471B	PPB		10 Apr 2019 10:58:31	U	SMPL
K2317-01MS - 1	SB-1_4.5-5.0MS	22679	3.7365	-7471B	PPB		10 Apr 2019 11:00:46	U	SMPL
K2317-01MSD - 1	SB-1_4.5-5.0MSD	22720	3.7434	-7471B	PPB		10 Apr 2019 11:03:02	U	SMPL
K2317-02 - 1	SB-2_3.5-4.0	391	-0.0183	-7471B	PPB		10 Apr 2019 11:05:18	U	SMPL
K2317-03 - 1	SB-3_3.5-4.0	371	-0.0217	-7471B	PPB		10 Apr 2019 11:07:35	U	SMPL
K2317-04 - 1	SB-4_3.5-4.0	309	-0.0322	-7471B	PPB		10 Apr 2019 11:09:52	U	SMPL
K2317-05 - 1	SB-5_4.0-4.5	2677	0.3668	-7471B	PPB		10 Apr 2019 11:12:09	U	SMPL
K2317-06 - 1	SB-6_2.5-3.0	362	-0.0232	-7471B	PPB		10 Apr 2019 11:14:26	U	SMPL
K2317-07 - 1	SB-7_3.0-3.5	6811	1.0632	-7471B	PPB		10 Apr 2019 11:16:41	U	SMPL
CCV62 - 1	CCV62	31028	5.1431	-7471B	PPB		10 Apr 2019 11:18:58	U	SMPL
CCB62 - 1	CCB62	261	-0.0402	-7471B	PPB		10 Apr 2019 11:21:13	U	SMPL
K2317-08 - 1	SB-8_2.5-3.0	22894	3.7728	-7471B	PPB		10 Apr 2019 11:23:29	U	SMPL
K2317-09 - 1	SB-9_3.5-4.0	589	0.0150	-7471B	PPB		10 Apr 2019 11:25:45	U	SMPL
K2317-10 - 1	SB-10_3.5-4.0	418	-0.0138	-7471B	PPB		10 Apr 2019 11:28:01	U	SMPL
K2317-11 - 1	SB-11_4.5-5.0	630	0.0219	-7471B	PPB		10 Apr 2019 11:30:16	U	SMPL
K2317-12 - 1	SB-12_3.5-4.0	1442	0.1587	-7471B	PPB		10 Apr 2019 11:32:32	U	SMPL
K2317-13 - 1	SB-13_5.5-6.0	883	0.0645	-7471B	PPB		10 Apr 2019 11:34:48	U	SMPL
K2317-14 - 1	COMP-1	632	0.0223	-7471B	PPB		10 Apr 2019 11:37:05	U	SMPL
K2317-15 - 1	COMP-2	617	0.0197	-7471B	PPB		10 Apr 2019 11:39:22	U	SMPL
K2317-16 - 1	COMP-3	613	0.0191	-7471B	PPB		10 Apr 2019 11:41:39	U	SMPL
K2334-02 - 1	P-6(0-10)	133928	22.4788	-7471B	PPB		10 Apr 2019 11:43:56	U	SMPL
CCV63 - 1	CCV63	30831	5.1099	-7471B	PPB		10 Apr 2019 11:46:12	U	SMPL
CCB63 - 1	CCB63	147	-0.0595	-7471B	PPB		10 Apr 2019 11:48:30	U	SMPL
K2338-01 - 1	TP-1-SHALLOW	1278	0.1311	-7471B	PPB		10 Apr 2019 11:50:48	U	SMPL
K2338-03 - 1	TP-1-DEEP	637	0.0231	-7471B	PPB		10 Apr 2019 11:53:04	U	SMPL
K2338-08 - 1	TP-2-SHALLOW	258	-0.0408	-7471B	PPB		10 Apr 2019 11:55:20	U	SMPL
K2338-10 - 1	TP-2-DEEP	558	0.0098	-7471B	PPB		10 Apr 2019 11:57:36	U	SMPL
K2338-12 - 1	TP-3-SHALLOW	4634	0.6965	-7471B	PPB		10 Apr 2019 11:59:53	U	SMPL
PB118759BL - 1	PBS	38	-0.0778	-7471B	PPB		10 Apr 2019 12:02:09	U	SMPL
PB118759BS - 1	LCSS	22247	3.6638	-7471B	PPB		10 Apr 2019 12:04:26	U	SMPL
K2200-01 - 1	EO-01-040919-A	1851	0.2276	-7471B	PPB		10 Apr 2019 12:06:43	U	SMPL
K2200-03 - 1	EO-01-040919-B	3567	0.5167	-7471B	PPB		10 Apr 2019 12:09:01	U	SMPL
K2200-05 - 1	EO-01-040919-C	3411	0.4904	-7471B	PPB		10 Apr 2019 12:11:17	U	SMPL
CCV64 - 1	CCV64	31104	5.1559	-7471B	PPB		10 Apr 2019 12:13:33	U	SMPL
CCB64 - 1	CCB64	214	-0.0482	-7471B	PPB		10 Apr 2019 12:15:49	U	SMPL
K2203-05 - 1	SU-04-040919-A	2607	0.3550	-7471B	PPB		10 Apr 2019 12:18:06	U	SMPL
K2203-07 - 1	SU-04-040919-B	1644	0.1927	-7471B	PPB		10 Apr 2019 12:20:23	U	SMPL
K2342-03 - 1	TP-11-S	119854	20.1077	-7471B	PPB		10 Apr 2019 12:22:39	U	SMPL
K2342-07 - 1	TP-11-D	136002	22.8282	-7471B	PPB		10 Apr 2019 12:24:55	U	SMPL
K2342-11 - 1	TP-12-S	89669	15.0224	-7471B	PPB		10 Apr 2019 12:27:14	U	SMPL
K2342-15 - 1	TP-12-D	22602	3.7236	-7471B	PPB		10 Apr 2019 12:29:30	U	SMPL
K2343-03 - 1	TP-7-S	1184	0.1153	-7471B	PPB		10 Apr 2019 12:31:49	U	SMPL
K2343-07 - 1	TP-7-D	467	-0.0055	-7471B	PPB		10 Apr 2019 12:33:53	U	SMPL
K2344-01 - 1	WOOD-1	479	-0.0035	-7471B	PPB		10 Apr 2019 12:36:09	U	SMPL
K2344-04 - 1	WOOD-2	575	0.0127	-7471B	PPB		10 Apr 2019 12:38:25	U	SMPL
CCV65 - 1	CCV65	31041	5.1453	-7471B	PPB		10 Apr 2019 12:40:41	U	SMPL

LB101982 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 10 Apr 2019 09:06:18

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
CCB65 - 1	CCB65	185	-0.0531	-7471B	PPB		10 Apr 2019 12:45:42	U	SMPL
K2345-01 - 1	TP-1	6722	1.0482	-7471B	PPB		10 Apr 2019 12:47:58	U	SMPL
K2345-04 - 1	TP-2	1738	0.2086	-7471B	PPB		10 Apr 2019 12:50:15	U	SMPL
K2345-07 - 1	TP-3	6716	1.0472	-7471B	PPB		10 Apr 2019 12:52:31	U	SMPL
K2345-10 - 1	TP-4	13164	2.1335	-7471B	PPB		10 Apr 2019 12:54:47	U	SMPL
K2346-01 - 1	VNJ-229	1978	0.2490	-7471B	PPB		10 Apr 2019 12:57:03	U	SMPL
K2346-01DUP - 1	VNJ-229DUP	2008	0.2541	-7471B	PPB		10 Apr 2019 12:59:20	U	SMPL
K2346-01MS - 1	VNJ-229MS	24292	4.0083	-7471B	PPB		10 Apr 2019 13:01:36	U	SMPL
K2346-01MSD - 1	VNJ-229MSD	24437	4.0327	-7471B	PPB		10 Apr 2019 13:03:53	U	SMPL
K2198-01LX5 - 1		237	-0.0443	-7471B	PPB		10 Apr 2019 13:06:11	U	SMPL
K2198-01A - 1		24542	4.0504	-7471B	PPB		10 Apr 2019 13:08:28	U	SMPL
CCV66 - 1	CCV66	30651	5.0796	-7471B	PPB		10 Apr 2019 13:10:45	U	SMPL
CCB66 - 1	CCB66	229	-0.0456	-7471B	PPB		10 Apr 2019 13:13:02	U	SMPL
K2317-01LX5 - 1		57	-0.0746	-7471B	PPB		10 Apr 2019 13:15:19	U	SMPL
K2317-01A - 1		22734	3.7458	-7471B	PPB		10 Apr 2019 13:17:35	U	SMPL
K2346-01LX5 - 1		307	-0.0325	-7471B	PPB		10 Apr 2019 13:19:52	U	SMPL
K2346-01A - 1		24311	4.0115	-7471B	PPB		10 Apr 2019 13:22:08	U	SMPL
K2316-01DLX4 - 1		26872	4.4429	-7471B	PPB		10 Apr 2019 13:24:25	U	SMPL
K2316-02DLX4 - 1		18638	3.0557	-7471B	PPB		10 Apr 2019 13:26:41	U	SMPL
K2316-03DLX4 - 1		26390	4.3617	-7471B	PPB		10 Apr 2019 13:28:58	U	SMPL
K2316-04DLX4 - 1		23849	3.9336	-7471B	PPB		10 Apr 2019 13:31:15	U	SMPL
K2316-05DLX2 - 1		31635	5.2454	-7471B	PPB		10 Apr 2019 13:33:31	U	SMPL
K2316-06DLX2 - 1		38180	6.3480	-7471B	PPB		10 Apr 2019 13:35:49	U	SMPL
CCV67 - 1	CCV67	31310	5.1906	-7471B	PPB		10 Apr 2019 13:38:06	U	SMPL
CCB67 - 1	CCB67	100	-0.0674	-7471B	PPB		10 Apr 2019 13:40:25	U	SMPL
K2316-07DLX2 - 1		31310	5.1906	-7471B	PPB		10 Apr 2019 13:42:42	U	SMPL
K2316-08DLX4 - 1		24303	4.0101	-7471B	PPB		10 Apr 2019 13:44:58	U	SMPL
K2316-09DLX4 - 1		18852	3.0918	-7471B	PPB		10 Apr 2019 13:47:15	U	SMPL
K2334-02DLX4 - 1		33798	5.6098	-7471B	PPB		10 Apr 2019 13:49:32	U	SMPL
K2342-03DLX4 - 1		30096	4.9861	-7471B	PPB		10 Apr 2019 13:51:49	U	SMPL
K2342-07DL4 - 1		34709	5.7632	-7471B	PPB		10 Apr 2019 13:54:05	U	SMPL
K2342-11DLX4 - 1		22294	3.6717	-7471B	PPB		10 Apr 2019 13:56:21	U	SMPL
CCV68 - 1	CCV68	31592	5.2381	-7471B	PPB		10 Apr 2019 13:58:39	U	SMPL
CCB68 - 1	CCB68	49	-0.0760	-7471B	PPB		10 Apr 2019 14:00:56	U	SMPL