



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB101991

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	04/10/2019	12:25
2	CAL2	1	Water	NA	NA	20.1	7.00	04/10/2019	12:26
3	CAL3	1	Water	NA	NA	20.1	10.03	04/10/2019	12:27
4	ICV	1	Water	NA	NA	20.2	7.01	04/10/2019	12:30
5	CCV1	1	Water	NA	NA	20.1	2.01	04/10/2019	12:31
6	K2351-01	1	Water	NA	NA	21.4	4.17	04/10/2019	12:33
7	K2351-01DUP	1	Water	NA	NA	21.5	4.18	04/10/2019	12:34
8	CCV2	1	Water	NA	NA	20.2	12.01	04/10/2019	12:35

1B101991

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K2351

WorkList ID : 124646

Date : 04-10-2019 12:14:44

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/04/2019	Water	K2351-01	pH	Cool 4 deg C	STER11	K11	POOL-WATER	04/04/2019	9040C

Date/Time 04.10.19 12.15
 Received by: Jn
 Relinquished by: cn

Date/Time 04.10.19 13:00
 Received by: cn
 Relinquished by: Jn