

LB102004


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

42043



A= 0.0000e+000

B= 2.3813e-004

C= -2.6538e-002

Rho= 0.9999738

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.012	0.012	161	0.000	161				
0.2	0.200	0.202	0.002	960	0.0 %	960				
2.5	2.500	2.493	-0.007	10579	0.0 %	10579				
5.0	5.000	4.960	-0.040	20941	0.0 %	20941				
7.5	7.500	7.548	0.048	31807	0.0 %	31807				
10.0	10.000	9.985	-0.015	42043	0.0 %	42043				

LB102004 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 10 Apr 2019 15:12:58

Sample ID	Extended ID	μ Abs	Conc	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50-0	161	-	0.00007470A	PPB	10 Apr 2019 15:23:35	S	Std
0.2 - 1	80-2	960	-	0.20007470A	PPB	10 Apr 2019 15:25:43	S	Std
2.5 - 1	52-5	10579	-	2.50007470A	PPB	10 Apr 2019 15:27:51	S	Std
5.0 - 1	55-0	20941	-	5.00007470A	PPB	10 Apr 2019 15:29:58	S	Std
7.5 - 1	57-5	31807	-	7.50007470A	PPB	10 Apr 2019 15:32:06	S	Std
10.0 - 1	510-0	42043	-	10.00007470A	PPB	10 Apr 2019 15:34:13	S	Std
ICV93 - 1	ICV93	16936	4.0065	-7470A	PPB	10 Apr 2019 15:39:13	U	SMPL
ICB93 - 1	ICB93	122	0.0025	-7470A	PPB	10 Apr 2019 15:41:20	U	SMPL
CCV69 - 1	CCV69	20957	4.9641	-7470A	PPB	10 Apr 2019 15:43:26	U	SMPL
CCB69 - 1	CCB69	27	-0.0201	-7470A	PPB	10 Apr 2019 15:45:32	U	SMPL
CRA - 1	CRA	932	0.1954	-7470A	PPB	10 Apr 2019 15:47:38	U	SMPL
HighStd - 1	HighStd	42318	10.0509	-7470A	PPB	10 Apr 2019 15:49:43	U	SMPL
ChkStd - 1	ChkStd	28707	6.8096	-7470A	PPB	10 Apr 2019 15:51:49	U	SMPL
PB118739BL - 1	PBW	109	-0.0006	-7470A	PPB	10 Apr 2019 15:53:57	U	SMPL
PB118739BS - 1	LCSW	17772	4.2056	-7470A	PPB	10 Apr 2019 15:56:04	U	SMPL
K2294-04 - 1	TP-14-S	322	0.0501	-7470A	PPB	10 Apr 2019 15:58:11	U	SMPL
K2294-04DUP - 1	TP-14-SDUP	351	0.0570	-7470A	PPB	10 Apr 2019 16:00:18	U	SMPL
K2294-04MS - 1	TP-14-SMS	17572	4.1580	-7470A	PPB	10 Apr 2019 16:02:26	U	SMPL
K2294-04MSD - 1	TP-14-SMSD	17397	4.1163	-7470A	PPB	10 Apr 2019 16:04:34	U	SMPL
K2294-08 - 1	TP-14-D	201	0.0213	-7470A	PPB	10 Apr 2019 16:06:43	U	SMPL
CCV70 - 1	CCV70	20387	4.8283	-7470A	PPB	10 Apr 2019 16:08:49	U	SMPL
CCB70 - 1	CCB70	57	-0.0130	-7470A	PPB	10 Apr 2019 16:10:55	U	SMPL
K2298-02 - 1	WALL-PAINT	297	0.0442	-7470A	PPB	10 Apr 2019 16:13:02	U	SMPL
K2298-04 - 1	FLOOR-PAINT	414	0.0720	-7470A	PPB	10 Apr 2019 16:15:08	U	SMPL
K2298-06 - 1	CELLING-PAINT	406	0.0701	-7470A	PPB	10 Apr 2019 16:17:14	U	SMPL
K2309-02 - 1	HA-040519	405	0.0699	-7470A	PPB	10 Apr 2019 16:19:22	U	SMPL
PB118740BL - 1	PBW	136	0.0058	-7470A	PPB	10 Apr 2019 16:21:29	U	SMPL
PB118740BS - 1	LCSW	16650	3.9384	-7470A	PPB	10 Apr 2019 16:23:36	U	SMPL
K2335-01 - 1	SV30755L	1164	0.2507	-7470A	PPB	10 Apr 2019 16:25:43	U	SMPL
K2337-01 - 1	WS040819	406	0.0701	-7470A	PPB	10 Apr 2019 16:27:51	U	SMPL
K2337-01DUP - 1	WS040819DUP	247	0.0323	-7470A	PPB	10 Apr 2019 16:29:59	U	SMPL
K2337-01MS - 1	WS040819MS	18841	4.4602	-7470A	PPB	10 Apr 2019 16:32:07	U	SMPL
CCV71 - 1	CCV71	20499	4.8550	-7470A	PPB	10 Apr 2019 16:34:14	U	SMPL
CCB71 - 1	CCB71	39	-0.0173	-7470A	PPB	10 Apr 2019 16:36:21	U	SMPL
K2337-01MSD - 1	WS040819MSD	19017	4.5021	-7470A	PPB	10 Apr 2019 16:38:27	U	SMPL
K2338-05 - 1	TP-1-WATER	2115	0.4771	-7470A	PPB	10 Apr 2019 16:40:33	U	SMPL
K2342-17 - 1	TP-12	41002	9.7375	-7470A	PPB	10 Apr 2019 16:42:40	U	SMPL
PB118742BL - 1	PBW	65	-0.0111	-7470A	PPB	10 Apr 2019 16:44:46	U	SMPL
PB118742BS - 1	LCSW	17373	4.1106	-7470A	PPB	10 Apr 2019 16:46:53	U	SMPL
K2337-02 - 1	WS040819	241	0.0309	-7470A	PPB	10 Apr 2019 16:49:00	U	SMPL
K2337-02DUP - 1	WS040819DUP	240	0.0306	-7470A	PPB	10 Apr 2019 16:51:07	U	SMPL
K2337-02MS - 1	WS040819MS	17500	4.1408	-7470A	PPB	10 Apr 2019 16:53:15	U	SMPL
K2337-02MSD - 1	WS040819MSD	17711	4.1911	-7470A	PPB	10 Apr 2019 16:55:23	U	SMPL
K2338-06 - 1	TP-1-WATER	135	0.0056	-7470A	PPB	10 Apr 2019 16:57:29	U	SMPL
CCV72 - 1	CCV72	21588	5.1143	-7470A	PPB	10 Apr 2019 16:59:36	U	SMPL
CCB72 - 1	CCB72	84	-0.0065	-7470A	PPB	10 Apr 2019 17:01:43	U	SMPL
PB118743BL - 1	PBW	206	0.0225	-7470A	PPB	10 Apr 2019 17:03:50	U	SMPL
PB118743BS - 1	LCSW	16655	3.9396	-7470A	PPB	10 Apr 2019 17:05:56	U	SMPL
K2198-02 - 1	OK-01-040819	161	0.0118	-7470A	PPB	10 Apr 2019 17:08:02	U	SMPL
K2198-02DUP - 1	OK-01-040819DUP	191	0.0189	-7470A	PPB	10 Apr 2019 17:10:09	U	SMPL
K2198-02MS - 1	OK-01-040819MS	17069	4.0382	-7470A	PPB	10 Apr 2019 17:12:16	U	SMPL
K2198-02MSD - 1	OK-01-040819MSD	17061	4.0363	-7470A	PPB	10 Apr 2019 17:14:24	U	SMPL
K2199-06 - 1	OR-03-040819-A	115	0.0008	-7470A	PPB	10 Apr 2019 17:16:32	U	SMPL
K2199-08 - 1	OR-03-040819-B	168	0.0135	-7470A	PPB	10 Apr 2019 17:18:39	U	SMPL
K2337-04 - 1	ODS-040819	155	0.0104	-7470A	PPB	10 Apr 2019 17:20:47	U	SMPL
K2338-02 - 1	TP-1-SHALLOW	225	0.0270	-7470A	PPB	10 Apr 2019 17:22:54	U	SMPL
CCV73 - 1	CCV73	22101	5.2365	-7470A	PPB	10 Apr 2019 17:25:01	U	SMPL
CCB73 - 1	CCB73	91	-0.0049	-7470A	PPB	10 Apr 2019 17:27:07	U	SMPL
K2338-04 - 1	TP-1-DEEP	129	0.0042	-7470A	PPB	10 Apr 2019 17:29:14	U	SMPL
K2338-09 - 1	TP-2-SHALLOW	150	0.0092	-7470A	PPB	10 Apr 2019 17:31:20	U	SMPL
K2338-11 - 1	TP-2-DEEP	145	0.0080	-7470A	PPB	10 Apr 2019 17:33:27	U	SMPL
K2338-13 - 1	TP-3-SHALLOW	171	0.0142	-7470A	PPB	10 Apr 2019 17:35:34	U	SMPL
PB118744BL - 1	PBW	145	0.0080	-7470A	PPB	10 Apr 2019 17:37:40	U	SMPL
PB118744BS - 1	LCSW	16735	3.9587	-7470A	PPB	10 Apr 2019 17:39:47	U	SMPL
K2316-11 - 1	SPLP-LECHATE-AMND-CS-1	423	0.0742	-7470A	PPB	10 Apr 2019 17:41:55	U	SMPL
K2316-11DUP - 1	SPLP-LECHATE-AMND-CS-1DUP	358	0.0587	-7470A	PPB	10 Apr 2019 17:44:03	U	SMPL
K2316-11MS - 1	SPLP-LECHATE-AMND-CS-1MS	17454	4.1299	-7470A	PPB	10 Apr 2019 17:46:11	U	SMPL
K2316-11MSD - 1	SPLP-LECHATE-AMND-CS-1MSD	17098	4.0451	-7470A	PPB	10 Apr 2019 17:48:18	U	SMPL
CCV74 - 1	CCV74	21944	5.1991	-7470A	PPB	10 Apr 2019 17:50:25	U	SMPL
CCB74 - 1	CCB74	88	-0.0056	-7470A	PPB	10 Apr 2019 17:52:32	U	SMPL
K2316-12 - 1	SPLP-LECHATE-AMND-CS-2	319	0.0494	-7470A	PPB	10 Apr 2019 17:54:40	U	SMPL
K2316-13 - 1	SPLP-LECHATE-AMND-CS-3	283	0.0409	-7470A	PPB	10 Apr 2019 17:56:47	U	SMPL
K2316-14 - 1	SPLP-LECHATE-AMND-CS-4	333	0.0528	-7470A	PPB	10 Apr 2019 17:58:55	U	SMPL
K2316-15 - 1	SPLP-LECHATE-AMND-CS-5	303	0.0456	-7470A	PPB	10 Apr 2019 18:01:02	U	SMPL
PB118739TB - 1	PB118699TB	170	0.0139	-7470A	PPB	10 Apr 2019 18:03:09	U	SMPL
PB118742TB - 1	PB118705TB	191	0.0189	-7470A	PPB	10 Apr 2019 18:05:16	U	SMPL
PB118743TB - 1	PB118699TB	168	0.0135	-7470A	PPB	10 Apr 2019 18:07:24	U	SMPL
PB118744TB - 1	PB118706TB	265	0.0366	-7470A	PPB	10 Apr 2019 18:09:31	U	SMPL
K2294-04LX5 - 1		58	-0.0127	-7470A	PPB	10 Apr 2019 18:11:39	U	SMPL
K2294-04A - 1		17771	4.2054	-7470A	PPB	10 Apr 2019 18:13:47	U	SMPL
CCV75 - 1	PB118739 CCV75	20379	4.8264	-7470A	PPB	10 Apr 2019 18:15:54	U	SMPL
CCB75 - 1	CCB75	70	-0.0099	-7470A	PPB	10 Apr 2019 18:18:02	U	SMPL
K2337-01LX5 - 1		49	-0.0149	-7470A	PPB	10 Apr 2019 18:20:09	U	SMPL
K2337-01A - 1	RS 04-10-19	16713	3.9534	-7470A	PPB	10 Apr 2019 18:22:16	U	SMPL
K2337-02LX5 - 1		-42	-0.0365	-7470A	PPB	10 Apr 2019 18:24:23	U	SMPL
K2337-02A - 1		17438	4.1261	-7470A	PPB	10 Apr 2019 18:26:30	U	SMPL
K2198-02LX5 - 1		-18	-0.0308	-7470A	PPB	10 Apr 2019 18:28:38	U	SMPL
K2198-02A - 1		17082	4.0413	-7470A	PPB	10 Apr 2019 18:30:45	U	SMPL
K2316-11LX5 - 1		32	-0.0189	-7470A	PPB	10 Apr 2019 18:32:53	U	SMPL
K2316-11A - 1		17325	4.0991	-7470A	PPB	10 Apr 2019 18:35:01	U	SMPL
CCV76 - 1	CCV76	21532	5.1010	-7470A	PPB	10 Apr 2019 18:37:08	U	SMPL
CCB76 - 1	CCB76	6	-0.0251	-7470A	PPB	10 Apr 2019 18:39:17	U	SMPL