

03/02/008

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

Instrument ID : Konelab

4/10/2019 15:30

Test: Phenolics-

Sample Id	Result ^{mg/L}	Dil. 1 +	Response	Errors
ICV1	1.049	0.0	0.122	
ICB1	0.001	0.0	0.013	
CCV1	1.048	0.0	0.122	
CCB1	-0.004	0.0	0.013	
PB118787BL	0.002	0.0	0.013	
PB118787BS	1.047	0.0	0.122	
PB118786BL	0.001	0.0	0.013	
PB118786BS	1.047	0.0	0.122	
K2241-07	0.010	0.0	0.014	
K2241-08	0.095	0.0	0.023	
K2241-01	0.008	0.0	0.014	
K2241-02	0.090	0.0	0.022	
CCV2	1.046	0.0	0.122	
CCB2	-0.002	0.0	0.013	

N 14
Mean 0.388
SD 0.5106
CV% 131.47

Aquakem v. 7.2AQ1

Results from time period:

Wed Apr 10 14:56:41 2019

Wed Apr 10 15:27:00 2019

✓AS

Sample Id	Sam/Ctr/c/	Test short nar	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.051	mg/l	4/10/2019 14:56:41	
0.1PPM	A	Phenolics-	P	0.1001	mg/l	4/10/2019 14:56:42	
0.25PPM	A	Phenolics-	P	0.2496	mg/l	4/10/2019 14:56:43	
0.50PPM	A	Phenolics-	P	0.5016	mg/l	4/10/2019 14:56:44	
1.0PPM	A	Phenolics-	P	0.9963	mg/l	4/10/2019 14:56:45	
2.0PPM	A	Phenolics-	P	2.0015	mg/l	4/10/2019 14:56:46	
ICV1	S	Phenolics-	P	1.0487	mg/l	4/10/2019 15:21:05	
ICB1	S	Phenolics-	P	0.0015	mg/l	4/10/2019 15:21:06	
CCV1	S	Phenolics-	P	1.0477	mg/l	4/10/2019 15:21:07	
CCB1	S	Phenolics-	P	-0.0043	mg/l	4/10/2019 15:21:08	
PB118787BL	S	Phenolics-	P	0.002	mg/l	4/10/2019 15:21:09	
PB118787BS	S	Phenolics-	P	1.0466	mg/l	4/10/2019 15:21:10	
PB118786BL	S	Phenolics-	P	0.0014	mg/l	4/10/2019 15:21:11	
PB118786BS	S	Phenolics-	P	1.0474	mg/l	4/10/2019 15:21:12	
K2241-07	S	Phenolics-	P	0.0097	mg/l	4/10/2019 15:21:13	
K2241-08	S	Phenolics-	P	0.0951	mg/l	4/10/2019 15:21:15	
K2241-01	S	Phenolics-	P	0.0076	mg/l	4/10/2019 15:26:55	
K2241-02	S	Phenolics-	P	0.0901	mg/l	4/10/2019 15:26:58	
CCV2	S	Phenolics-	P	1.0455	mg/l	4/10/2019 15:26:59	
CCB2	S	Phenolics-	P	-0.002	mg/l	4/10/2019 15:27:00	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

Instrument ID : Konelab

4/10/2019 14:57


Test Phenolics-

Accepted 4/10/2019 14:57

Factor 9.615
Bias 0.013

Coeff. of det. 0.999993

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.018	0.0510	0.0500	
2	Phenol-2	0.023	0.1001	0.1000	
3	Phenol-2	0.039	0.2496	0.2500	
4	Phenol-2	0.065	0.5016	0.5000	
5	Phenol-2	0.117	0.9963	1.0000	
6	Phenol-2	0.221	2.0015	2.0000	