



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 4/11/2019

OVENTEMP IN Celsius(°C): 108
Time IN: 17:05
In Date: 04/10/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 2.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 102
Time OUT: 08:10
Out Date: 04/11/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-1
Thermometer ID: %SOLIDS-OVEN

QC:LB102012

Lab ID	Client Sample ID	Dish#	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish + Dry Sample Wt(g) (C)	% Solid
K2355-01	TP-9-S1	1	1.14	9.73	9.16	93.4
K2355-02	TP-9-S2	2	1.13	9.95	8.91	88.2
K2355-03	TP-9-S	3	1.15	9.98	9.35	92.9
K2356-01	S-1	4	1.17	9.62	9.02	92.9
K2356-02	S-1A	5	1.15	9.7	9.09	92.9
K2357-01	CS-041019	6	1.00	2.00	2.00	100.0
K2358-01	TP-5-S1	7	1.16	9.65	8.77	89.6
K2358-02	TP-5-S2	8	1.16	9.68	8.21	82.7
K2358-03	TP-5	9	1.14	9.61	8.62	88.3
K2358-05	TP-9-S1	10	1.16	9.73	8.82	89.4
K2358-06	TP-9-S2	11	1.13	9.84	8.57	85.4
K2358-07	TP-9	12	1.16	9.8	8.8	88.4
K2358-09	TP-7-S1	13	1.14	9.95	7.51	72.3
K2358-10	TP-7-S2	14	1.15	9.59	7.09	70.4
K2358-11	TP-7-S	15	1.15	9.7	7.00	68.4
K2358-13	TP-7-D1	16	1.12	9.67	8.74	89.1
K2358-14	TP-7-D2	17	1.12	9.97	9.00	89.0
K2358-15	TP-7-D	18	1.13	9.98	9.09	89.9
K2360-01	FK-BPM-01-E2-E2A-F2-F2A	19	1.14	9.72	9.13	93.1
K2360-02	FK-BPM-01-E2-E2A-F2-F2A	20	1.16	9.61	9.03	93.1
K2361-01	FK-BPM-01-E1-E1A-F1-F1A	21	1.12	9.63	9.04	93.1
K2361-02	FK-BPM-01-E1-E1A-F1-F1A	22	1.17	9.55	8.98	93.2

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$